

### Initial and Continuing Calibration Blank Summary

**Client:** Portal Partners Tri-Venture  
**Project:** Amtrak Sawtooth Bridges 2025

**SDG No.:** Q1744  
**RunNo.:** LB135331

| Analyte                                       | Units | Result   | Acceptance Limits | Conc Qual | MDL    | RDL  | Analysis Date |
|-----------------------------------------------|-------|----------|-------------------|-----------|--------|------|---------------|
| Sample ID: <b>ICB</b><br>Hexavalent Chromium  | mg/L  | < 0.0050 | 0.0050            | U         | 0.0029 | 0.01 | 04/07/2025    |
| Sample ID: <b>CCB1</b><br>Hexavalent Chromium | mg/L  | < 0.0050 | 0.0050            | U         | 0.0029 | 0.01 | 04/07/2025    |
| Sample ID: <b>CCB2</b><br>Hexavalent Chromium | mg/L  | < 0.0050 | 0.0050            | U         | 0.0029 | 0.01 | 04/07/2025    |
| Sample ID: <b>CCB3</b><br>Hexavalent Chromium | mg/L  | < 0.0050 | 0.0050            | U         | 0.0029 | 0.01 | 04/07/2025    |

### Initial and Continuing Calibration Blank Summary

**Client:** Portal Partners Tri-Venture  
**Project:** Amtrak Sawtooth Bridges 2025

**SDG No.:** Q1744  
**RunNo.:** LB135341

| Analyte                                    | Units | Result   | Acceptance Limits | Conc Qual | MDL     | RDL   | Analysis Date |
|--------------------------------------------|-------|----------|-------------------|-----------|---------|-------|---------------|
| Sample ID: <b>ICB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 04/08/2025    |
| Sample ID: <b>CCB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 04/08/2025    |
| Sample ID: <b>CCB2</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 04/08/2025    |
| Sample ID: <b>CCB3</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 04/08/2025    |