



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	Q1744			
Contract:	PORT06	Lab Code:	CHEM		Case No.:	Q1744	SAS No.:	Q1744	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB82	Mercury	0.20	+/-0.20	U	0.20	CV	04/09/2025	11:36	LB135354

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB49	Mercury	0.20	+/-0.20	U	0.20	CV	04/09/2025	11:45	LB135354
CCB50	Mercury	0.20	+/-0.20	U	0.20	CV	04/09/2025	12:20	LB135354
CCB51	Mercury	0.20	+/-0.20	U	0.20	CV	04/09/2025	12:48	LB135354
CCB52	Mercury	0.20	+/-0.20	U	0.20	CV	04/09/2025	14:04	LB135354
CCB53	Mercury	0.20	+/-0.20	U	0.20	CV	04/09/2025	14:21	LB135354

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1744							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1744	SAS No.: Q1744						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	10:29	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	10:29	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	10:29	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	10:29	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	10:29	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1744							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1744	SAS No.: Q1744						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	11:21	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	11:21	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	11:21	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	11:21	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	11:21	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
CCB02	Aluminum	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	12:12	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	12:12	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	12:12	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	12:12	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	12:12	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
CCB03	Aluminum	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	13:04	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:04	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:04	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	13:04	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	13:04	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
CCB04	Aluminum	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	13:55	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:55	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:55	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	13:55	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	13:55	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
CCB05	Aluminum	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	14:49	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	14:49	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	14:49	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	14:49	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	14:49	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
CCB06	Aluminum	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	15:48	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	15:48	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	15:48	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1744							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1744	SAS No.: Q1744						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	15:48	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	15:48	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
CCB07	Aluminum	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	17:00	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	17:00	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	17:00	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	17:00	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	17:00	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
CCB08	Aluminum	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	18:00	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:00	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:00	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	18:00	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	18:00	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
CCB09	Aluminum	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	18:23	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:23	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:23	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	18:23	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	18:23	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	Q1744			
Contract:	PORT06	Lab Code:	CHEM	Case No.:	Q1744	SAS No.:	Q1744		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373