

LB135365

7470A

INSTRUMENT ID: CV1

μ Abs.:

47882

Linear

A= 0.0000e+000

B= 2.1114e-004

C= 2.8382e-002

Rho= 0.9996275

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.046	0.046	82	0.000	82					
0.2	0.200	0.201	0.001	817	0.0 %	817	0				
2.5	2.500	2.508	0.008	11745	0.0 %	11745					
5.0	5.000	5.006	0.006	23577	0.0 %	23577					
7.5	7.500	7.301	-0.199	34445	0.0 %	34445					
10.0	10.000	10.138	0.138	47882	0.0 %	47882					

10031

Method: 7470A

Operator: Admin

LB135365 INSTRUMENT ID : CV1

Date of Analysis: 10 Apr 2025 09:47:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc Method	Units	Date	Type	Type
0.0 - 1	50	82	-	0.00007470A	PPB	10 Apr 2025	12:41:43	S Std
0.2 - 1	50.2	817	-	0.20007470A	PPB	10 Apr 2025	12:44:01	S Std
2.5 - 1	52.5	11745	-	2.50007470A	PPB	10 Apr 2025	12:46:18	S Std
5.0 - 1	55	23577	-	5.00007470A	PPB	10 Apr 2025	12:48:35	S Std
7.5 - 1	57.5	34445	-	7.50007470A	PPB	10 Apr 2025	12:50:53	S Std
10.0 - 1	59.1	47882	-	10.00007470A	PPB	10 Apr 2025	12:53:10	S Std
ICV83 - 1	ICV83	19091	4.0592	-7470A	PPB	10 Apr 2025	13:01:29	U SMPL
ICB83 - 1	ICB83	-223	-0.0187	-7470A	PPB	10 Apr 2025	13:03:44	U SMPL
CCV53 - 1	CCV53	23474	4.9846	-7470A	PPB	10 Apr 2025	13:06:02	U SMPL
CCB53 - 1	CCB53	-314	-0.0379	-7470A	PPB	10 Apr 2025	13:08:17	U SMPL
CRA - 1	CRA	774	0.1918	-7470A	PPB	10 Apr 2025	13:10:35	U SMPL
HighStd - 1	HighStd	46900	9.9306	-7470A	PPB	10 Apr 2025	13:12:49	U SMPL
ChkStd - 1	ChkStd	30395	6.4458	-7470A	PPB	10 Apr 2025	13:15:04	U SMPL
PB167539BL - 1	PBW	-370	-0.0497	-7470A	PPB	10 Apr 2025	13:17:23	U SMPL
PB167539BS - 1	LCSW	17884	3.8043	-7470A	PPB	10 Apr 2025	13:22:27	U SMPL
Q1743-04 - 1	TP-16	-200	-0.0138	-7470A	PPB	10 Apr 2025	13:24:43	U SMPL
Q1743-04DUP - 1	TP-16DUP	-53	0.0172	-7470A	PPB	10 Apr 2025	13:27:01	U SMPL
Q1743-04MS - 1	TP-16MS	17208	3.6616	-7470A	PPB	10 Apr 2025	13:31:56	U SMPL
Q1743-04MSD - 1	TP-16MSD	16676	3.5493	-7470A	PPB	10 Apr 2025	13:34:12	U SMPL
Q1744-02 - 1	B-158-SB01	-186	-0.0109	-7470A	PPB	10 Apr 2025	13:36:30	U SMPL
CCV54 - 1	CCV54	23308	4.9495	-7470A	PPB	10 Apr 2025	13:38:47	U SMPL
CCB54 - 1	CCB54	-214	-0.0168	-7470A	PPB	10 Apr 2025	13:41:02	U SMPL
Q1744-04 - 1	B-158-SB02	11	0.0307	-7470A	PPB	10 Apr 2025	13:43:20	U SMPL
Q1745-04 - 1	IB-6A-WC	73	0.0438	-7470A	PPB	10 Apr 2025	13:45:36	U SMPL
Q1745-12 - 1	IB-6.5-WC	-33	0.0214	-7470A	PPB	10 Apr 2025	13:47:51	U SMPL
Q1746-02 - 1	B-149-SB01	-36	0.0208	-7470A	PPB	10 Apr 2025	13:50:06	U SMPL
Q1746-04 - 1	B-149-SB02	16	0.0318	-7470A	PPB	10 Apr 2025	13:52:23	U SMPL
Q1748-04 - 1	IB-1.5-WC	47	0.0383	-7470A	PPB	10 Apr 2025	13:54:39	U SMPL
Q1748-08 - 1	IB-2A-WC	56	0.0402	-7470A	PPB	10 Apr 2025	13:56:55	U SMPL
Q1749-04 - 1	TP-14	29	0.0345	-7470A	PPB	10 Apr 2025	13:59:13	U SMPL
Q1752-02 - 1	TP-1	69	0.0430	-7470A	PPB	10 Apr 2025	14:01:30	U SMPL
Q1752-04 - 1	TP-2	22	0.0330	-7470A	PPB	10 Apr 2025	14:03:46	U SMPL
CCV55 - 1	CCV55	22979	4.8801	-7470A	PPB	10 Apr 2025	14:06:02	U SMPL
CCB55 - 1	CCB55	-313	-0.0377	-7470A	PPB	10 Apr 2025	14:08:18	U SMPL
Q1752-06 - 1	TP-4	-49	0.0180	-7470A	PPB	10 Apr 2025	14:10:35	U SMPL
Q1752-08 - 1	TP-3	12	0.0309	-7470A	PPB	10 Apr 2025	14:12:51	U SMPL
Q1753-04 - 1	WC-1	20	0.0326	-7470A	PPB	10 Apr 2025	14:15:07	U SMPL
Q1753-08 - 1	WC-2	48	0.0385	-7470A	PPB	10 Apr 2025	14:17:23	U SMPL
Q1754-02 - 1	TP-1	89	0.0472	-7470A	PPB	10 Apr 2025	14:19:38	U SMPL
Q1754-04 - 1	TP-1-CONCRETE	45	0.0379	-7470A	PPB	10 Apr 2025	14:21:55	U SMPL
PB167540BL - 1	PBW	17	0.0320	-7470A	PPB	10 Apr 2025	14:24:12	U SMPL
PB167540BS - 1	LCSW	19042	4.0488	-7470A	PPB	10 Apr 2025	14:26:29	U SMPL
Q1730-02 - 1	OU4-VSL-15-040325	2759	0.6109	-7470A	PPB	10 Apr 2025	14:28:46	U SMPL
Q1730-04 - 1	OU4-VSL-16-040325	157	0.0615	-7470A	PPB	10 Apr 2025	14:31:04	U SMPL
CCV56 - 1	CCV56	23272	4.9419	-7470A	PPB	10 Apr 2025	14:36:06	U SMPL
CCB56 - 1	CCB56	-261	-0.0267	-7470A	PPB	10 Apr 2025	14:38:21	U SMPL
Q1730-06 - 1	OU4-VSL-17-040325	818	0.2011	-7470A	PPB	10 Apr 2025	14:40:39	U SMPL
Q1730-08 - 1	OU4-PCS-TC-21-040325	134	0.0567	-7470A	PPB	10 Apr 2025	14:42:54	U SMPL
Q1730-10 - 1	OU4-PCS-TC-22-040325	163	0.0628	-7470A	PPB	10 Apr 2025	14:45:10	U SMPL
Q1730-12 - 1	OU4-PCS-TC-23-040325	147	0.0594	-7470A	PPB	10 Apr 2025	14:47:25	U SMPL
Q1730-14 - 1	OU4-PCS-TC-24-040325	0	0.0284	-7470A	PPB	10 Apr 2025	14:49:42	U SMPL
Q1730-16 - 1	OU4-PCS-TC-25-040325	115	0.0527	-7470A	PPB	10 Apr 2025	14:51:58	U SMPL
Q1730-18 - 1	OU4-PCS-TC-26-040325	112	0.0520	-7470A	PPB	10 Apr 2025	14:54:14	U SMPL
Q1730-20 - 1	OU4-CF-15-040325	97	0.0489	-7470A	PPB	10 Apr 2025	14:56:31	U SMPL
Q1730-20DUP - 1	OU4-CF-15-040325DUP	125	0.0548	-7470A	PPB	10 Apr 2025	14:58:48	U SMPL
Q1730-20MS - 1	OU4-CF-15-040325MS	18894	4.0176	-7470A	PPB	10 Apr 2025	15:03:49	U SMPL
CCV57 - 1	CCV57	22666	4.8140	-7470A	PPB	10 Apr 2025	15:06:04	U SMPL
CCB57 - 1	CCB57	-257	-0.0259	-7470A	PPB	10 Apr 2025	15:08:22	U SMPL
Q1730-20MSD - 1	OU4-CF-15-040325MSD	18537	3.9422	-7470A	PPB	10 Apr 2025	15:10:40	U SMPL
PB167517TB - 1	PB167517TB	-223	-0.0187	-7470A	PPB	10 Apr 2025	15:12:55	U SMPL
PB167519TB - 1	PB167519TB	17	0.0320	-7470A	PPB	10 Apr 2025	15:15:13	U SMPL
Q1743-04LX5 - 1	Q1743-04LX5	18	0.0322	-7470A	PPB	10 Apr 2025	15:17:28	U SMPL
Q1743-04A - 1	Q1743-04A	16402	3.4914	-7470A	PPB	10 Apr 2025	15:19:44	U SMPL
Q1730-20LX5 - 1	Q1730-20LX5	45	0.0379	-7470A	PPB	10 Apr 2025	15:22:01	U SMPL
Q1730-20A - 1	Q1730-20A	16145	3.4372	-7470A	PPB	10 Apr 2025	15:24:18	U SMPL
CCV58 - 1	CCV58	23589	5.0089	-7470A	PPB	10 Apr 2025	15:26:35	U SMPL
CCB58 - 1	CCB58	-313	-0.0377	-7470A	PPB	10 Apr 2025	15:28:54	U SMPL