

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV83	Mercury	4.06	4.0	102	90 - 110	CV	04/10/2025	13:01	LB135365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV53	Mercury	4.98	5.0	100	90 - 110	CV	04/10/2025	13:06	LB135365
CCV54	Mercury	4.95	5.0	99	90 - 110	CV	04/10/2025	13:38	LB135365
CCV55	Mercury	4.88	5.0	98	90 - 110	CV	04/10/2025	14:06	LB135365
CCV56	Mercury	4.94	5.0	99	90 - 110	CV	04/10/2025	14:36	LB135365
CCV57	Mercury	4.81	5.0	96	90 - 110	CV	04/10/2025	15:06	LB135365
CCV58	Mercury	5.01	5.0	100	90 - 110	CV	04/10/2025	15:26	LB135365

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	985	1000	98	90 - 110	P	04/10/2025	10:09	LB135373
	Barium	534	520	103	90 - 110	P	04/10/2025	10:09	LB135373
	Cadmium	483	510	95	90 - 110	P	04/10/2025	10:09	LB135373
	Chromium	505	520	97	90 - 110	P	04/10/2025	10:09	LB135373
	Lead	951	1000	95	90 - 110	P	04/10/2025	10:09	LB135373
	Selenium	1030	1000	103	90 - 110	P	04/10/2025	10:09	LB135373
	Silver	245	250	98	90 - 110	P	04/10/2025	10:09	LB135373



- 2a -

Client:	<u>Portal Partners Tri-Venture</u>	SDG No.:	<u>Q1744</u>	
Contract:	<u>PORT06</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1744</u> SAS No.: <u>Q1744</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
LLICV01	Arsenic	18.5	20.0	92	80 - 120	P	04/10/2025	10:25	LB135373
	Barium	85.5	100	86	80 - 120	P	04/10/2025	10:25	LB135373
	Cadmium	5.62	6.0	94	80 - 120	P	04/10/2025	10:25	LB135373
	Chromium	9.56	10.0	96	80 - 120	P	04/10/2025	10:25	LB135373
	Lead	9.94	12.0	83	80 - 120	P	04/10/2025	10:25	LB135373
	Selenium	18.9	20.0	94	80 - 120	P	04/10/2025	10:25	LB135373
	Silver	10.5	10.0	104	80 - 120	P	04/10/2025	10:25	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4750	5000	95	90 - 110	P	04/10/2025	11:00	LB135373
	Barium	9770	10000	98	90 - 110	P	04/10/2025	11:00	LB135373
	Cadmium	2340	2500	94	90 - 110	P	04/10/2025	11:00	LB135373
	Chromium	1010	1000	101	90 - 110	P	04/10/2025	11:00	LB135373
	Lead	5240	5000	105	90 - 110	P	04/10/2025	11:00	LB135373
	Selenium	4820	5000	96	90 - 110	P	04/10/2025	11:00	LB135373
	Silver	1250	1250	100	90 - 110	P	04/10/2025	11:00	LB135373
CCV02	Arsenic	4800	5000	96	90 - 110	P	04/10/2025	12:08	LB135373
	Barium	9680	10000	97	90 - 110	P	04/10/2025	12:08	LB135373
	Cadmium	2350	2500	94	90 - 110	P	04/10/2025	12:08	LB135373
	Chromium	1020	1000	102	90 - 110	P	04/10/2025	12:08	LB135373
	Lead	4720	5000	94	90 - 110	P	04/10/2025	12:08	LB135373
	Selenium	4860	5000	97	90 - 110	P	04/10/2025	12:08	LB135373
	Silver	1260	1250	101	90 - 110	P	04/10/2025	12:08	LB135373
CCV03	Arsenic	4890	5000	98	90 - 110	P	04/10/2025	12:59	LB135373
	Barium	10100	10000	101	90 - 110	P	04/10/2025	12:59	LB135373
	Cadmium	2360	2500	94	90 - 110	P	04/10/2025	12:59	LB135373
	Chromium	1020	1000	102	90 - 110	P	04/10/2025	12:59	LB135373
	Lead	4740	5000	95	90 - 110	P	04/10/2025	12:59	LB135373
	Selenium	4990	5000	100	90 - 110	P	04/10/2025	12:59	LB135373
	Silver	1270	1250	101	90 - 110	P	04/10/2025	12:59	LB135373
CCV04	Arsenic	4880	5000	98	90 - 110	P	04/10/2025	13:51	LB135373
	Barium	9600	10000	96	90 - 110	P	04/10/2025	13:51	LB135373
	Cadmium	2380	2500	95	90 - 110	P	04/10/2025	13:51	LB135373
	Chromium	1020	1000	102	90 - 110	P	04/10/2025	13:51	LB135373
	Lead	4760	5000	95	90 - 110	P	04/10/2025	13:51	LB135373
	Selenium	4960	5000	99	90 - 110	P	04/10/2025	13:51	LB135373
	Silver	1270	1250	102	90 - 110	P	04/10/2025	13:51	LB135373
CCV05	Arsenic	4820	5000	96	90 - 110	P	04/10/2025	14:44	LB135373
	Barium	9730	10000	97	90 - 110	P	04/10/2025	14:44	LB135373
	Cadmium	2350	2500	94	90 - 110	P	04/10/2025	14:44	LB135373
	Chromium	1020	1000	102	90 - 110	P	04/10/2025	14:44	LB135373
	Lead	4720	5000	94	90 - 110	P	04/10/2025	14:44	LB135373
	Selenium	4910	5000	98	90 - 110	P	04/10/2025	14:44	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1270	1250	101	90 - 110	P	04/10/2025	14:44	LB135373
CCV06	Arsenic	4780	5000	96	90 - 110	P	04/10/2025	15:44	LB135373
	Barium	9640	10000	96	90 - 110	P	04/10/2025	15:44	LB135373
	Cadmium	2340	2500	93	90 - 110	P	04/10/2025	15:44	LB135373
	Chromium	1010	1000	101	90 - 110	P	04/10/2025	15:44	LB135373
	Lead	4690	5000	94	90 - 110	P	04/10/2025	15:44	LB135373
	Selenium	4850	5000	97	90 - 110	P	04/10/2025	15:44	LB135373
	Silver	1260	1250	100	90 - 110	P	04/10/2025	15:44	LB135373
CCV07	Arsenic	4920	5000	98	90 - 110	P	04/10/2025	16:56	LB135373
	Barium	9570	10000	96	90 - 110	P	04/10/2025	16:56	LB135373
	Cadmium	2350	2500	94	90 - 110	P	04/10/2025	16:56	LB135373
	Chromium	1020	1000	102	90 - 110	P	04/10/2025	16:56	LB135373
	Lead	4740	5000	95	90 - 110	P	04/10/2025	16:56	LB135373
	Selenium	5020	5000	100	90 - 110	P	04/10/2025	16:56	LB135373
	Silver	1260	1250	101	90 - 110	P	04/10/2025	16:56	LB135373
CCV08	Arsenic	4890	5000	98	90 - 110	P	04/10/2025	17:56	LB135373
	Barium	9470	10000	95	90 - 110	P	04/10/2025	17:56	LB135373
	Cadmium	2360	2500	94	90 - 110	P	04/10/2025	17:56	LB135373
	Chromium	1030	1000	103	90 - 110	P	04/10/2025	17:56	LB135373
	Lead	4740	5000	95	90 - 110	P	04/10/2025	17:56	LB135373
	Selenium	4980	5000	100	90 - 110	P	04/10/2025	17:56	LB135373
	Silver	1260	1250	101	90 - 110	P	04/10/2025	17:56	LB135373
CCV09	Arsenic	4850	5000	97	90 - 110	P	04/10/2025	18:19	LB135373
	Barium	9730	10000	97	90 - 110	P	04/10/2025	18:19	LB135373
	Cadmium	2350	2500	94	90 - 110	P	04/10/2025	18:19	LB135373
	Chromium	1000	1000	100	90 - 110	P	04/10/2025	18:19	LB135373
	Lead	4720	5000	94	90 - 110	P	04/10/2025	18:19	LB135373
	Selenium	4930	5000	99	90 - 110	P	04/10/2025	18:19	LB135373
	Silver	1240	1250	99	90 - 110	P	04/10/2025	18:19	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1744

Contract: PORT06 Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1060	1000	106	90 - 110	P	04/14/2025	13:56	LB135424
	Barium	534	520	103	90 - 110	P	04/14/2025	13:56	LB135424
	Cadmium	507	510	99	90 - 110	P	04/14/2025	13:56	LB135424
	Chromium	507	520	98	90 - 110	P	04/14/2025	13:56	LB135424
	Lead	1000	1000	100	90 - 110	P	04/14/2025	13:56	LB135424
	Selenium	1050	1000	105	90 - 110	P	04/14/2025	13:56	LB135424
	Silver	246	250	98	90 - 110	P	04/14/2025	13:56	LB135424

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1744

Contract: PORT06 Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	22.8	20.0	114	80 - 120	P	04/14/2025	14:01	LB135424
	Barium	96.1	100	96	80 - 120	P	04/14/2025	14:01	LB135424
	Cadmium	6.12	6.0	102	80 - 120	P	04/14/2025	14:01	LB135424
	Chromium	9.81	10.0	98	80 - 120	P	04/14/2025	14:01	LB135424
	Lead	12.4	12.0	103	80 - 120	P	04/14/2025	14:01	LB135424
	Selenium	19.0	20.0	95	80 - 120	P	04/14/2025	14:01	LB135424
	Silver	10.0	10.0	100	80 - 120	P	04/14/2025	14:01	LB135424

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4910	5000	98	90 - 110	P	04/14/2025	14:31	LB135424
	Barium	9640	10000	96	90 - 110	P	04/14/2025	14:31	LB135424
	Cadmium	2460	2500	98	90 - 110	P	04/14/2025	14:31	LB135424
	Chromium	1000	1000	100	90 - 110	P	04/14/2025	14:31	LB135424
	Lead	4940	5000	99	90 - 110	P	04/14/2025	14:31	LB135424
	Selenium	4930	5000	98	90 - 110	P	04/14/2025	14:31	LB135424
	Silver	1250	1250	100	90 - 110	P	04/14/2025	14:31	LB135424
CCV02	Arsenic	4930	5000	99	90 - 110	P	04/14/2025	15:22	LB135424
	Barium	9580	10000	96	90 - 110	P	04/14/2025	15:22	LB135424
	Cadmium	2460	2500	98	90 - 110	P	04/14/2025	15:22	LB135424
	Chromium	995	1000	100	90 - 110	P	04/14/2025	15:22	LB135424
	Lead	4930	5000	98	90 - 110	P	04/14/2025	15:22	LB135424
	Selenium	4970	5000	99	90 - 110	P	04/14/2025	15:22	LB135424
	Silver	1230	1250	98	90 - 110	P	04/14/2025	15:22	LB135424
CCV03	Arsenic	4890	5000	98	90 - 110	P	04/14/2025	16:28	LB135424
	Barium	9870	10000	99	90 - 110	P	04/14/2025	16:28	LB135424
	Cadmium	2470	2500	99	90 - 110	P	04/14/2025	16:28	LB135424
	Chromium	983	1000	98	90 - 110	P	04/14/2025	16:28	LB135424
	Lead	4930	5000	99	90 - 110	P	04/14/2025	16:28	LB135424
	Selenium	4900	5000	98	90 - 110	P	04/14/2025	16:28	LB135424
	Silver	1230	1250	99	90 - 110	P	04/14/2025	16:28	LB135424
CCV04	Arsenic	4920	5000	98	90 - 110	P	04/14/2025	17:18	LB135424
	Barium	9580	10000	96	90 - 110	P	04/14/2025	17:18	LB135424
	Cadmium	2480	2500	99	90 - 110	P	04/14/2025	17:18	LB135424
	Chromium	1000	1000	100	90 - 110	P	04/14/2025	17:18	LB135424
	Lead	4940	5000	99	90 - 110	P	04/14/2025	17:18	LB135424
	Selenium	4950	5000	99	90 - 110	P	04/14/2025	17:18	LB135424
	Silver	1240	1250	99	90 - 110	P	04/14/2025	17:18	LB135424
CCV05	Arsenic	4860	5000	97	90 - 110	P	04/14/2025	18:16	LB135424
	Barium	9570	10000	96	90 - 110	P	04/14/2025	18:16	LB135424
	Cadmium	2470	2500	99	90 - 110	P	04/14/2025	18:16	LB135424
	Chromium	986	1000	99	90 - 110	P	04/14/2025	18:16	LB135424
	Lead	4920	5000	98	90 - 110	P	04/14/2025	18:16	LB135424
	Selenium	4910	5000	98	90 - 110	P	04/14/2025	18:16	LB135424

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1744

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1230	1250	98	90 - 110	P	04/14/2025	18:16	LB135424
CCV06	Arsenic	4780	5000	96	90 - 110	P	04/14/2025	19:07	LB135424
	Barium	9740	10000	97	90 - 110	P	04/14/2025	19:07	LB135424
	Cadmium	2460	2500	98	90 - 110	P	04/14/2025	19:07	LB135424
	Chromium	957	1000	96	90 - 110	P	04/14/2025	19:07	LB135424
	Lead	4870	5000	98	90 - 110	P	04/14/2025	19:07	LB135424
	Selenium	4820	5000	96	90 - 110	P	04/14/2025	19:07	LB135424
	Silver	1190	1250	95	90 - 110	P	04/14/2025	19:07	LB135424
CCV07	Arsenic	4740	5000	95	90 - 110	P	04/14/2025	19:57	LB135424
	Barium	9610	10000	96	90 - 110	P	04/14/2025	19:57	LB135424
	Cadmium	2430	2500	97	90 - 110	P	04/14/2025	19:57	LB135424
	Chromium	974	1000	97	90 - 110	P	04/14/2025	19:57	LB135424
	Lead	4820	5000	96	90 - 110	P	04/14/2025	19:57	LB135424
	Selenium	4770	5000	95	90 - 110	P	04/14/2025	19:57	LB135424
	Silver	1200	1250	96	90 - 110	P	04/14/2025	19:57	LB135424
CCV08	Arsenic	4810	5000	96	90 - 110	P	04/14/2025	20:48	LB135424
	Barium	9830	10000	98	90 - 110	P	04/14/2025	20:48	LB135424
	Cadmium	2440	2500	98	90 - 110	P	04/14/2025	20:48	LB135424
	Chromium	955	1000	96	90 - 110	P	04/14/2025	20:48	LB135424
	Lead	4850	5000	97	90 - 110	P	04/14/2025	20:48	LB135424
	Selenium	4860	5000	97	90 - 110	P	04/14/2025	20:48	LB135424
	Silver	1190	1250	96	90 - 110	P	04/14/2025	20:48	LB135424
CCV09	Arsenic	4770	5000	95	90 - 110	P	04/14/2025	21:38	LB135424
	Barium	9610	10000	96	90 - 110	P	04/14/2025	21:38	LB135424
	Cadmium	2430	2500	97	90 - 110	P	04/14/2025	21:38	LB135424
	Chromium	973	1000	97	90 - 110	P	04/14/2025	21:38	LB135424
	Lead	4840	5000	97	90 - 110	P	04/14/2025	21:38	LB135424
	Selenium	4800	5000	96	90 - 110	P	04/14/2025	21:38	LB135424
	Silver	1220	1250	97	90 - 110	P	04/14/2025	21:38	LB135424
CCV10	Arsenic	4830	5000	97	90 - 110	P	04/14/2025	22:28	LB135424
	Barium	9670	10000	97	90 - 110	P	04/14/2025	22:28	LB135424
	Cadmium	2450	2500	98	90 - 110	P	04/14/2025	22:28	LB135424
	Chromium	977	1000	98	90 - 110	P	04/14/2025	22:28	LB135424
	Lead	4870	5000	97	90 - 110	P	04/14/2025	22:28	LB135424

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1744
Contract: PORT06 Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Selenium	4880	5000	98	90 - 110	P	04/14/2025	22:28	LB135424
	Silver	1220	1250	98	90 - 110	P	04/14/2025	22:28	LB135424
CCV11	Arsenic	4800	5000	96	90 - 110	P	04/14/2025	22:40	LB135424
	Barium	9830	10000	98	90 - 110	P	04/14/2025	22:40	LB135424
	Cadmium	2450	2500	98	90 - 110	P	04/14/2025	22:40	LB135424
	Chromium	958	1000	96	90 - 110	P	04/14/2025	22:40	LB135424
	Lead	4850	5000	97	90 - 110	P	04/14/2025	22:40	LB135424
	Selenium	4850	5000	97	90 - 110	P	04/14/2025	22:40	LB135424
	Silver	1210	1250	97	90 - 110	P	04/14/2025	22:40	LB135424
CCV12	Arsenic	4760	5000	95	90 - 110	P	04/14/2025	23:06	LB135424
	Barium	9900	10000	99	90 - 110	P	04/14/2025	23:06	LB135424
	Cadmium	2430	2500	97	90 - 110	P	04/14/2025	23:06	LB135424
	Chromium	965	1000	96	90 - 110	P	04/14/2025	23:06	LB135424
	Lead	4830	5000	97	90 - 110	P	04/14/2025	23:06	LB135424
	Selenium	4800	5000	96	90 - 110	P	04/14/2025	23:06	LB135424
	Silver	1200	1250	96	90 - 110	P	04/14/2025	23:06	LB135424