



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	18.9	20.0	95	40 - 160	P	04/10/2025	10:33	LB135373
	Barium	85.4	100	85	40 - 160	P	04/10/2025	10:33	LB135373
	Cadmium	5.51	6.0	92	40 - 160	P	04/10/2025	10:33	LB135373
	Chromium	9.42	10.0	94	40 - 160	P	04/10/2025	10:33	LB135373
	Lead	11.6	12.0	97	40 - 160	P	04/10/2025	10:33	LB135373
	Selenium	18.1	20.0	91	40 - 160	P	04/10/2025	10:33	LB135373
	Silver	10.4	10.0	104	40 - 160	P	04/10/2025	10:33	LB135373
CRA	Mercury	0.19	0.2	96	40 - 160	CV	04/10/2025	13:10	LB135365
CRI01	Arsenic	22.3	20.0	112	40 - 160	P	04/14/2025	14:09	LB135424
	Barium	95.6	100	96	40 - 160	P	04/14/2025	14:09	LB135424
	Cadmium	6.08	6.0	101	40 - 160	P	04/14/2025	14:09	LB135424
	Chromium	9.67	10.0	97	40 - 160	P	04/14/2025	14:09	LB135424
	Lead	12.2	12.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Selenium	19.6	20.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Silver	9.88	10.0	99	40 - 160	P	04/14/2025	14:09	LB135424