



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	Q1744				
Contract:	PORT06	Lab Code:	CHEM		Case No.:	Q1744	SAS No.:	Q1744		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB83	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	13:03	LB135365	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB53	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	13:08	LB135365
CCB54	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	13:41	LB135365
CCB55	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	14:08	LB135365
CCB56	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	14:38	LB135365
CCB57	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	15:08	LB135365
CCB58	Mercury	0.20	+/-0.20	U	0.20	CV	04/10/2025	15:28	LB135365

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** Q1744
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	10:29	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	10:29	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1744							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1744	SAS No.: Q1744						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	11:21	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	11:21	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	12:12	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	12:12	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:04	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	13:04	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:55	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	13:55	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	14:49	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	14:49	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	15:48	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** Q1744
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1744 **SAS No.:** Q1744

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	15:48	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	17:00	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	17:00	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:00	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	18:00	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
CCB09	Barium	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:23	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	18:23	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Portal Partners Tri-Venture</u>		SDG No.: <u>Q1744</u>							
Contract: <u>PORT06</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1744</u>		SAS No.: <u>Q1744</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	14:05	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	14:05	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	14:05	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	14:05	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	14:05	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	14:05	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	14:05	LB135424

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1744							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1744	SAS No.: Q1744						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	14:35	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	14:35	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	14:35	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	14:35	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	14:35	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	14:35	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	14:35	LB135424
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	15:26	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	15:26	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	15:26	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	15:26	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	15:26	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	15:26	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	15:26	LB135424
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	16:32	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	16:32	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	16:32	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	16:32	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	16:32	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	16:32	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	16:32	LB135424
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	17:22	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	17:22	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	17:22	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	17:22	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	17:22	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	17:22	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	17:22	LB135424
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	18:20	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	18:20	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	18:20	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	18:20	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	18:20	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	18:20	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	18:20	LB135424
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	19:11	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	19:11	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	19:11	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	19:11	LB135424

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: Q1744							
Contract: PORT06	Lab Code: CHEM	Case No.: Q1744	SAS No.: Q1744						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	19:11	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	19:11	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	19:11	LB135424
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	20:01	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	20:01	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	20:01	LB135424
CCB08	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	20:01	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	20:01	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	20:01	LB135424
CCB09	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	20:01	LB135424
	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	20:52	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	20:52	LB135424
CCB10	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	20:52	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	20:52	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	20:52	LB135424
CCB11	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	20:52	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	20:52	LB135424
	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	21:42	LB135424
CCB12	Barium	100	+/-100	U	100	P	04/14/2025	21:42	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	21:42	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	21:42	LB135424
CCB13	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	21:42	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	21:42	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	21:42	LB135424
CCB14	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	22:32	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	22:32	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	22:32	LB135424
CCB15	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	22:32	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	22:32	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	22:32	LB135424
CCB16	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	22:32	LB135424
	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	22:44	LB135424
	Barium	100	+/-100	U	100	P	04/14/2025	22:44	LB135424
CCB17	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	22:44	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	22:44	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	22:44	LB135424
CCB18	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	22:44	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	22:44	LB135424
	Arsenic	20.0	+/-20.0	U	20.0	P	04/14/2025	23:11	LB135424

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Portal Partners Tri-Venture</u>		SDG No.: <u>Q1744</u>							
Contract: <u>PORT06</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1744</u>	SAS No.: <u>Q1744</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Barium	100	+/-100	U	100	P	04/14/2025	23:11	LB135424
	Cadmium	6.00	+/-6.00	U	6.00	P	04/14/2025	23:11	LB135424
	Chromium	10.0	+/-10.0	U	10.0	P	04/14/2025	23:11	LB135424
	Lead	12.0	+/-12.0	U	12.0	P	04/14/2025	23:11	LB135424
	Selenium	20.0	+/-20.0	U	20.0	P	04/14/2025	23:11	LB135424
	Silver	10.0	+/-10.0	U	10.0	P	04/14/2025	23:11	LB135424