

Continuing Calibration Report for SequenceID : FD040825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049254.D

Aromatic C16-C21	13883380.000	80.000	8.818	13.108	173542.250	167701.469	-3.483
Aromatic C21-C36	25180825.000	180.000	13.109	18.525	139893.472	138140.429	-1.269
Aromatic EPH	56448595.000	360.000	4.441	18.525	156801.653	152129.398	-3.071
Aromatic C10-C12	6617916.000	40.000	4.441	6.187	165447.900	156792.089	-5.521
Aromatic C12-C16	10766474.000	60.000	6.188	8.817	179441.233	170225.083	-5.414