

Continuing Calibration Report for SequenceID : FE040825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053257.D

Aliphatic C9-C12	8399438.000	60.000	3.139	6.782	139990.633	156032.227	10.281
Aliphatic C12-C16	5671262.000	40.000	6.783	10.234	141781.550	153631.377	7.713
Aliphatic C16-C21	8041295.000	60.000	10.235	13.613	134021.583	144337.201	7.147
Aliphatic C21-C28	10233470.000	80.000	13.614	17.285	127918.375	134347.450	4.785
Aliphatic C28-C40	15856211.000	120.000	17.286	22.199	132135.092	124256.607	-6.340
Aliphatic EPH	48201676.000	360.000	3.139	22.199	133893.544	138405.582	3.260