

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1744 **SAS No :** Q1744 **SDG No:** Q1744
Sample No : PB167513BSD **Datafile:** FE053260.D
Client ID : PB167513BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10.0	6.25	62		5.7 (40-140)	50
Aliphatic C12-C16	6.7	5.72	86		5.3 (40-140)	50
Aliphatic C16-C21	10.0	9.82	98		5.7 (40-140)	50
Aliphatic C21-C28	13.3	13.5	101		5.5 (40-140)	50
Aliphatic C28-C40	20.0	23.3	116		16.1 (40-140)	50