

Continuing Calibration Report for SequenceID : FC041125AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC068559.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 6096249.000  | 60.000  | 3.345  | 6.644  | 101604.150 | 105282.003 | 3.493  |
| Aliphatic C12-C16 | 3784483.000  | 40.000  | 6.645  | 10.043 | 94612.075  | 102322.575 | 7.535  |
| Aliphatic C16-C21 | 5304758.000  | 60.000  | 10.044 | 13.409 | 88412.633  | 98816.539  | 10.529 |
| Aliphatic C21-C28 | 6712207.000  | 80.000  | 13.410 | 17.073 | 83902.588  | 94185.337  | 10.918 |
| Aliphatic C28-C40 | 10067118.000 | 120.000 | 17.074 | 22.075 | 83892.650  | 89882.835  | 6.664  |
| Aliphatic EPH     | 31964815.000 | 360.000 | 3.345  | 22.075 | 88791.153  | 96276.618  | 7.775  |