

Continuing Calibration Report for SequenceID : FC041125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068571.D

Aliphatic C9-C12	6233414.000	60.000	3.345	6.644	103890.233	105282.003	1.322
Aliphatic C12-C16	3778079.000	40.000	6.645	10.043	94451.975	102322.575	7.692
Aliphatic C16-C21	5202338.000	60.000	10.044	13.409	86705.633	98816.539	12.256
Aliphatic C21-C28	6676284.000	80.000	13.410	17.073	83453.550	94185.337	11.394
Aliphatic C28-C40	10202260.000	120.000	17.074	22.075	85018.833	89882.835	5.411
Aliphatic EPH	32092375.000	360.000	3.345	22.075	89145.486	96276.618	7.407