

Continuing Calibration Report for SequenceID : FD041025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049268.D

Aromatic C10-C12	5439988.000	40.000	4.440	6.185	135999.700	156792.089	13.261
Aromatic C12-C16	8722016.000	60.000	6.186	8.813	145366.933	170225.083	14.603
Aromatic C16-C21	11673840.000	80.000	8.814	13.102	145923.000	167701.469	12.986
Aromatic C21-C36	23479622.000	180.000	13.103	18.519	130442.344	138140.429	5.573
Aromatic EPH	49315466.000	360.000	4.440	18.519	136987.406	152129.398	9.953