

Continuing Calibration Report for SequenceID : FD041025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049280.D

Aromatic C10-C12	5275482.000	40.000	4.440	6.185	131887.050	156792.089	15.884
Aromatic C12-C16	8494935.000	60.000	6.186	8.813	141582.250	170225.083	16.826
Aromatic C16-C21	11446135.000	80.000	8.814	13.102	143076.688	167701.469	14.684
Aromatic C21-C36	22930748.000	180.000	13.103	18.519	127393.044	138140.429	7.780
Aromatic EPH	48147300.000	360.000	4.440	18.519	133742.500	152129.398	12.086