

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1744 **SAS No :** Q1744 **SDG No:** Q1744
Sample No : PB167513BS **Datafile:** FC068561.D
Client ID : PB167513BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	5.72	57		(40-140)
Aliphatic C12-C16	6.7	5.43	82		(40-140)
Aliphatic C16-C21	10	9.70	97		(40-140)
Aliphatic C21-C28	13.3	13.1	98		(40-140)
Aliphatic C28-C40	20.0	19.1	96		(40-140)