



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1744

Order ID: Q1744

Client: Portal Partners Tri-Venture

Project ID: Amtrak Sawtooth Bridges 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A

B

C

D

E

F

G

H

I

J

K

L



SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/05/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/07/25
Client Sample ID:	B-158-SB01	SDG No.:	Q1744
Lab Sample ID:	Q1744-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071150.D	1	04/08/25 08:11	04/08/25 14:30	PB167507

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	18.9	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	18.9	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.9	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	18.9	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	18.9	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60	18.9	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	18.9	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	18.9	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	18.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	30.3	*	30 (32) - 150 (144)	151%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.3		30 (32) - 150 (175)	116%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/05/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/07/25
Client Sample ID:	B-158-SB02	SDG No.:	Q1744
Lab Sample ID:	Q1744-03	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	56
Sample Wt/Vol:	30.09	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071158.D	1	04/08/25 08:11	04/08/25 17:40	PB167507

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	7.00	U	7.00	30.3	ug/kg
11104-28-2	Aroclor-1221	7.20	U	7.20	30.3	ug/kg
11141-16-5	Aroclor-1232	6.60	U	6.60	30.3	ug/kg
53469-21-9	Aroclor-1242	7.10	U	7.10	30.3	ug/kg
12672-29-6	Aroclor-1248	10.5	U	10.5	30.3	ug/kg
11097-69-1	Aroclor-1254	5.70	U	5.70	30.3	ug/kg
37324-23-5	Aroclor-1262	8.90	U	8.90	30.3	ug/kg
11100-14-4	Aroclor-1268	6.40	U	6.40	30.3	ug/kg
11096-82-5	Aroclor-1260	5.80	U	5.80	30.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.0		30 (32) - 150 (144)	115%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.9		30 (32) - 150 (175)	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: **Q1744**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070916.D	PIBLK-PP070916.D	Tetrachloro-m-xylene	1	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	19.4	97		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.1	91		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.4	117		70 (60)	130 (140)
I.BLK-PP071140.D	PIBLK-PP071140.D	Tetrachloro-m-xylene	1	20	22.8	114		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.9	119		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	23.1	116		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.5	108		70 (60)	130 (140)
Q1742-01MS	TR-06-040725MS	Tetrachloro-m-xylene	1	20	18.7	94		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	23.0	115		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.9	100		30 (32)	150 (175)
Q1742-01MSD	TR-06-040725MSD	Tetrachloro-m-xylene	1	20	17.4	87		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.3	112		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.2	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.5	103		30 (32)	150 (175)
Q1744-01	B-158-SB01	Tetrachloro-m-xylene	1	20	24.8	124		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	23.3	116		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	30.3	151	*	30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.9	99		30 (32)	150 (175)
I.BLK-PP071155.D	PIBLK-PP071155.D	Tetrachloro-m-xylene	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.0	105		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.1	105		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.4	92		70 (60)	130 (140)
PB167507BL	PB167507BL	Tetrachloro-m-xylene	1	20	26.5	132		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	25.9	130		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	27.7	138		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	24.0	120		30 (32)	150 (175)
PB167507BS	PB167507BS	Tetrachloro-m-xylene	1	20	26.3	132		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	25.7	129		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	24.7	123		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.6	113		30 (32)	150 (175)
Q1744-03	B-158-SB02	Tetrachloro-m-xylene	1	20	22.8	114		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.9	84		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	23.0	115		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	15.4	77		30 (32)	150 (175)
I.BLK-PP071168.D	PIBLK-PP071168.D	Tetrachloro-m-xylene	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.0	105		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.8	99		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.1	100		70 (60)	130 (140)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1744

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

DataFile : PP071147.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	TR-06-040725MS											
Q1742-01MS	AR1016	189.9	0	176	ug/kg	93				40 (55)	140 (146)	
	AR1260	189.9	0	188	ug/kg	99				40 (31)	140 (146)	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1744

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

DataFile : PP071148.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	TR-06-040725MSD											
Q1742-01MSD	AR1016	189.7	0	173	ug/kg	91		2		40 (55)	140 (146)	30 (20)
	AR1260	189.7	0	177	ug/kg	93		6		40 (31)	140 (146)	30 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1744

Client: Portal Partners Tri-Venture

Analytical Method: **8082A**

Datafile : PP071157.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB167507BS	AR1016	166.6	181	ug/kg	109				40 (71)	140 (120)		
	AR1260	166.6	188	ug/kg	113				40 (65)	140 (130)		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167507BL

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: Q1744

SAS No.: Q1744 SDG NO.: Q1744

Lab Sample ID: PB167507BL

Lab File ID: PP071156.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 04/08/2025

Date Analyzed (1): 04/08/2025

Date Analyzed (2): 04/08/2025

Time Analyzed (1): 17:07

Time Analyzed (2): 17:07

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
TR-06-040725MS	Q1742-01MS	PP071147.D	04/08/2025	04/08/2025
TR-06-040725MSD	Q1742-01MSD	PP071148.D	04/08/2025	04/08/2025
B-158-SB01	Q1744-01	PP071150.D	04/08/2025	04/08/2025
PB167507BS	PB167507BS	PP071157.D	04/08/2025	04/08/2025
B-158-SB02	Q1744-03	PP071158.D	04/08/2025	04/08/2025

COMMENTS: _____



QC SAMPLE DATA

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	
Client Sample ID:	PB167507BL	SDG No.:	Q1744
Lab Sample ID:	PB167507BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.02 Units: g	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071156.D	1	04/08/25 08:11	04/08/25 17:07	PB167507

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	27.7		30 (32) - 150 (144)	138%	SPK: 20
2051-24-3	Decachlorobiphenyl	25.9		30 (32) - 150 (175)	130%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	03/27/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	03/27/25
Client Sample ID:	PIBLK-PP070916.D	SDG No.:	Q1744
Lab Sample ID:	I.BLK-PP070916.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070916.D	1		03/27/25	PP032725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		70 (60) - 130 (140)	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		70 (60) - 130 (140)	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/08/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/08/25
Client Sample ID:	PIBLK-PP071140.D	SDG No.:	Q1744
Lab Sample ID:	I.BLK-PP071140.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071140.D	1		04/08/25	PP040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.8		70 (60) - 130 (140)	114%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.5		70 (60) - 130 (140)	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/08/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/08/25
Client Sample ID:	PIBLK-PP071155.D	SDG No.:	Q1744
Lab Sample ID:	I.BLK-PP071155.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071155.D	1		04/08/25	PP040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.6		70 (60) - 130 (140)	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.4		70 (60) - 130 (140)	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/08/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/08/25
Client Sample ID:	PIBLK-PP071168.D	SDG No.:	Q1744
Lab Sample ID:	I.BLK-PP071168.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071168.D	1		04/08/25	PP040825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		70 (60) - 130 (140)	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.1		70 (60) - 130 (140)	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:		
Project:	Amtrak Sawtooth Bridges 2025		Date Received:		
Client Sample ID:	PB167507BS		SDG No.:	Q1744	
Lab Sample ID:	PB167507BS		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071157.D	1	04/08/25 08:11	04/08/25 17:23	PB167507

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	181		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	188		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.3		30 (32) - 150 (144)	132%	SPK: 20
2051-24-3	Decachlorobiphenyl	25.7		30 (32) - 150 (175)	129%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/04/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/04/25
Client Sample ID:	TR-06-040725MS	SDG No.:	Q1744
Lab Sample ID:	Q1742-01MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	87.6
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071147.D	1	04/08/25 08:11	04/08/25 13:41	PB167507

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	176		4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	188		3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		30 (32) - 150 (144)	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		30 (32) - 150 (175)	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	04/04/25
Project:	Amtrak Sawtooth Bridges 2025	Date Received:	04/04/25
Client Sample ID:	TR-06-040725MSD	SDG No.:	Q1744
Lab Sample ID:	Q1742-01MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	87.6
Sample Wt/Vol:	30.09 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071148.D	1	04/08/25 08:11	04/08/25 13:57	PB167507

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	173		4.50	19.3	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.3	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.3	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.3	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.3	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.3	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.3	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.3	ug/kg
11096-82-5	Aroclor-1260	177		3.70	19.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.2		30 (32) - 150 (144)	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.3		30 (32) - 150 (175)	112%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

A
B
C
D
E
F
G
H
I
J
K
L

Calibration Times: 10:30 17:53

LAB FILE ID: RT 1000 = PP070917.D RT 750 = PP070918.D
RT 500 = PP070919.D RT 250 = PP070920.D RT 050 = PP070921.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.23	10.23	10.23	10.23	10.23	10.23	10.13	10.33
Tetrachloro-m-xylene	4.52	4.51	4.51	4.52	4.51	4.51	4.41	4.61

A
B
C
D
E
F
G
H
I
J
K
L

GC Column: ZB-MR2 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.86	8.86	8.86	8.86	8.86	8.86	8.76	8.96
Tetrachloro-m-xylene	3.82	3.82	3.82	3.82	3.82	3.82	3.72	3.92

A
B
C
D
E
F
G
H
I
J
K
L

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM **Case No.:** Q1744 **SAS No.:** Q1744 **SDG NO.:** Q1744

Instrument ID: ECD_P **Calibration Date(s):** 03/27/2025 03/27/2025
Calibration Times: 10:30 17:53

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP070917.D</u>		CF 750 = <u>PP070918.D</u>			
CF 500 = <u>PP070919.D</u>		CF 250 = <u>PP070920.D</u>		CF 050 = <u>PP070921.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	47478014	49171687	50934060	54324668	52347880	50851262
Aroclor-1016-2	(2)	72483549	74568483	78811172	80433884	83971080	78053634
Aroclor-1016-3	(3)	43541298	44043953	46815864	48849960	52621940	47174603
Aroclor-1016-4	(4)	36661280	36602855	39014928	39212148	32287440	36755730
Aroclor-1016-5	(5)	32204227	33487635	35040682	36610940	32217200	33912137
Aroclor-1260-1	(1)	60625379	62238407	66823960	72729028	73347280	67152811
Aroclor-1260-2	(2)	90805769	95497569	97808710	102858984	93710780	96136362
Aroclor-1260-3	(3)	75946924	78957439	82970652	86114216	75601760	79918198
Aroclor-1260-4	(4)	67190656	75020355	81451428	78145636	91876440	78736903
Aroclor-1260-5	(5)	154592706	156108101	159960800	166857288	156911780	158886135
Decachlorobiphenyl		1019047430	1065730867	1116449000	1115108760	946747800	1052616771
Tetrachloro-m-xylene		1424837400	1461405720	1519248320	1595716440	1306667000	1461574976
Aroclor-1242-1	(1)	40990074	42746281	43432670	43984636	48732080	43977148
Aroclor-1242-2	(2)	64704598	63862937	67024876	68850388	57138840	64316328
Aroclor-1242-3	(3)	39349869	38048504	39702972	40627540	47468380	41039453
Aroclor-1242-4	(4)	30814559	31466187	32277086	32194644	25864220	30523339
Aroclor-1242-5	(5)	36006298	34981232	37334376	38560396	35669740	36510408
Decachlorobiphenyl		1036298130	1063489027	1106293120	1153452680	1056566000	1083219791
Tetrachloro-m-xylene		1443947330	1507747640	1537397120	1567165920	1326698400	1476591282
Aroclor-1248-1	(1)	31141042	33379292	33781538	45988760	37146900	36287506
Aroclor-1248-2	(2)	42399779	44979081	46192256	47122176	43999640	44938586
Aroclor-1248-3	(3)	48055172	50777464	51825052	50697180	51832600	50637494
Aroclor-1248-4	(4)	59878410	63471275	67039892	73429036	56919340	64147591
Aroclor-1248-5	(5)	56652876	60504408	64286140	68519384	54268440	60846250
Decachlorobiphenyl		1072017300	1124462173	1205136680	1167091680	1109917800	1135725127
Tetrachloro-m-xylene		1462181050	1544091280	1549039360	1559485200	1464932600	1515945898
Aroclor-1254-1	(1)	60157506	63067431	65471886	66806416	66447920	64390232
Aroclor-1254-2	(2)	92838442	97047215	101802392	105556104	108208740	101090579
Aroclor-1254-3	(3)	92334650	98055471	99416554	103841140	104991280	99727819
Aroclor-1254-4	(4)	84238161	87780547	86802510	88788836	85373700	86596751
Aroclor-1254-5	(5)	74433164	79038301	80854650	83958584	83470140	80350968
Decachlorobiphenyl		1095432410	1144692320	1168417220	1194330360	1035516600	1127677782
Tetrachloro-m-xylene		1501288360	1551275560	1575746420	1626904120	1301718600	1511386612
Aroclor-1268-1	(1)	226575551	237927453	239660032	247892192	225482800	235507606

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	190380276	198934901	200232346	204948924	182079900	195315269	4
Aroclor-1268-3	(3)	163648333	171737784	173955356	177246576	167979860	170913582	3
Aroclor-1268-4	(4)	70377008	73665913	73237284	74802952	61652780	70747187	7
Aroclor-1268-5	(5)	459553484	481598519	482514296	498456744	458756600	476175929	3
Decachlorobiphenyl		1956984350	2055290840	2095781900	2171413960	2007035400	2057301290	4
Tetrachloro-m-xylene		1523485290	1594639080	1607574540	1645012520	1440121400	1562166566	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Instrument ID: ECD_P

Calibration Date(s): 03/27/2025 03/27/2025

Calibration Times: 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP070917.D</u>		CF 750 = <u>PP070918.D</u>			
CF 500 = <u>PP070919.D</u>		CF 250 = <u>PP070920.D</u>		CF 050 = <u>PP070921.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	34208953	37661787	37955886	45294856	45319420	40088180
Aroclor-1016-2	(2)	49610574	53393839	54380778	65503040	66437080	57865062
Aroclor-1016-3	(3)	27000462	29732755	29849058	35508136	32526120	30923306
Aroclor-1016-4	(4)	21229955	23465415	23793788	28635320	24655080	24355912
Aroclor-1016-5	(5)	26939544	29893109	30154374	37292184	37326660	32321174
Aroclor-1260-1	(1)	45249076	46519276	48240522	56226688	59796020	51206316
Aroclor-1260-2	(2)	58475471	61260192	62904038	69183788	67145740	63793846
Aroclor-1260-3	(3)	50913622	52312629	53920638	58739540	56547700	54486826
Aroclor-1260-4	(4)	42366044	44395528	47488628	49955128	54926140	47826294
Aroclor-1260-5	(5)	113066934	116639771	125312810	127263128	140268800	124510289
Decachlorobiphenyl		793547960	846236520	865016160	753569280	682611000	788196184
Tetrachloro-m-xylene		926488500	1023245573	1036748080	1124812280	1071924800	1036643847
Aroclor-1242-1	(1)	30391529	31772724	34649408	38698720	35466820	34195840
Aroclor-1242-2	(2)	43879042	45075739	49434800	55503780	52395240	49257720
Aroclor-1242-3	(3)	24103124	24630729	26618958	29833472	28035100	26644277
Aroclor-1242-4	(4)	23086538	23159225	25581538	28618736	31445140	26378235
Aroclor-1242-5	(5)	28647762	29235153	31558738	32962404	32387440	30958299
Decachlorobiphenyl		794686010	779465320	775544600	851493360	677466800	775731218
Tetrachloro-m-xylene		994474400	1020857307	1119824860	1120535680	1106593200	1072457089
Aroclor-1248-1	(1)	23297586	26737011	28501250	29207540	27656520	27079981
Aroclor-1248-2	(2)	31071420	35918237	38274460	38702620	39645560	36722459
Aroclor-1248-3	(3)	32592570	37701093	39964120	40477712	42561380	38659375
Aroclor-1248-4	(4)	38114246	43411732	47181776	48244244	48753000	45141000
Aroclor-1248-5	(5)	37031094	41115012	43925318	46634268	44688040	42678746
Decachlorobiphenyl		763432400	822192813	827097300	806313080	732328400	790272799
Tetrachloro-m-xylene		969522160	1045815880	1138159760	1084071480	1169729800	1081459816
Aroclor-1254-1	(1)	58928147	61799583	66082486	73787844	69490960	66017804
Aroclor-1254-2	(2)	50145967	53786207	56793554	64034708	60730300	57098147
Aroclor-1254-3	(3)	77448562	83642240	87764778	99090376	86260880	86841367
Aroclor-1254-4	(4)	49998215	54614948	55966900	64174348	55747980	56100478
Aroclor-1254-5	(5)	65979008	71358709	71825896	82425644	69163880	72150627
Decachlorobiphenyl		783483810	796181933	910832540	850283840	759277600	820011945
Tetrachloro-m-xylene		1042595600	1091536387	1108578240	1195164040	1067355200	1101045893
Aroclor-1268-1	(1)	165356134	175541800	174753666	176934688	164423820	171402022

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	137801505	147739099	145758868	146778256	138336660	143282878	3
Aroclor-1268-3	(3)	120132746	127507232	124391202	127247744	116467620	123149309	4
Aroclor-1268-4	(4)	51096977	54505272	53344994	55689972	49428660	52813175	5
Aroclor-1268-5	(5)	357060145	376226307	370949858	366544744	338563340	361868879	4
Decachlorobiphenyl		1484820890	1579158933	1563503220	1620729960	1432317200	1536106041	5
Tetrachloro-m-xylene		1051289840	1118865693	1118283520	1183542080	1260888000	1146573827	7

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Instrument ID: ECD_P Date(s) Analyzed: 03/27/2025 03/27/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.72	4.62	4.82	17353700
		2	4.80	4.70	4.90	12952900
		3	4.88	4.78	4.98	41717200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.88	4.78	4.98	33288800
		2	5.41	5.31	5.51	15910300
		3	5.69	5.59	5.79	34270400
		4	5.85	5.75	5.95	15707600
		5	5.94	5.84	6.04	10640400
Aroclor-1262	500	1	8.09	7.99	8.19	106300000
		2	8.41	8.31	8.51	198383000
		3	8.73	8.63	8.83	131742000
		4	8.82	8.72	8.92	95614800
		5	9.47	9.37	9.57	65066400

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Instrument ID: ECD_P Date(s) Analyzed: 03/27/2025 03/27/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.03	3.93	4.13	15243600
		2	4.11	4.01	4.21	11436300
		3	4.19	4.09	4.29	34089000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.19	4.09	4.29	27986800
		2	4.92	4.82	5.02	28061200
		3	5.10	5.00	5.20	14889600
		4	5.19	5.09	5.29	12804400
		5	5.36	5.26	5.46	14174800
Aroclor-1262	500	1	6.95	6.85	7.05	85563600
		2	7.21	7.11	7.31	68464400
		3	7.73	7.63	7.83	60723600
		4	7.79	7.69	7.89	99406800
		5	8.29	8.19	8.39	46871600

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 09:37 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.68	5.66	5.56	5.76	-0.02
Aroclor-1016-2	(2)	5.70	5.69	5.59	5.79	-0.01
Aroclor-1016-3	(3)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1016-4	(4)	5.86	5.85	5.75	5.95	-0.01
Aroclor-1016-5	(5)	6.15	6.14	6.04	6.24	-0.01
Aroclor-1260-1	(1)	7.27	7.26	7.16	7.36	-0.01
Aroclor-1260-2	(2)	7.53	7.51	7.41	7.61	-0.02
Aroclor-1260-3	(3)	7.89	7.87	7.77	7.97	-0.02
Aroclor-1260-4	(4)	8.11	8.09	7.99	8.19	-0.02
Aroclor-1260-5	(5)	8.43	8.41	8.31	8.51	-0.02
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	-0.01
Decachlorobiphenyl		10.26	10.23	10.13	10.33	-0.03

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 09:37 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-3	(3)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-4	(4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5	(5)	5.36	5.36	5.26	5.46	0.00
Aroclor-1260-1	(1)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-2	(2)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-3	(3)	6.74	6.73	6.63	6.83	-0.01
Aroclor-1260-4	(4)	7.21	7.21	7.11	7.31	0.00
Aroclor-1260-5	(5)	7.45	7.45	7.35	7.55	0.00
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071136.D Time Analyzed: 09:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.679	5.563	5.763	504.230	500.000	0.8
Aroclor-1016-2	5.700	5.585	5.785	472.510	500.000	-5.5
Aroclor-1016-3	5.763	5.648	5.848	488.820	500.000	-2.2
Aroclor-1016-4	5.861	5.745	5.945	525.740	500.000	5.1
Aroclor-1016-5	6.154	6.038	6.238	539.860	500.000	8.0
Aroclor-1260-1	7.273	7.156	7.356	518.030	500.000	3.6
Aroclor-1260-2	7.527	7.410	7.610	536.780	500.000	7.4
Aroclor-1260-3	7.885	7.768	7.968	503.360	500.000	0.7
Aroclor-1260-4	8.109	7.992	8.192	501.370	500.000	0.3
Aroclor-1260-5	8.431	8.311	8.511	506.740	500.000	1.3
Decachlorobiphenyl	10.255	10.125	10.325	51.060	50.000	2.1
Tetrachloro-m-xylene	4.524	4.412	4.612	51.340	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071136.D Time Analyzed: 09:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.904	4.803	5.003	491.030	500.000	-1.8
Aroclor-1016-2	4.923	4.822	5.022	483.510	500.000	-3.3
Aroclor-1016-3	5.101	4.999	5.199	522.710	500.000	4.5
Aroclor-1016-4	5.142	5.041	5.241	542.480	500.000	8.5
Aroclor-1016-5	5.357	5.255	5.455	530.130	500.000	6.0
Aroclor-1260-1	6.393	6.292	6.492	547.620	500.000	9.5
Aroclor-1260-2	6.582	6.480	6.680	534.460	500.000	6.9
Aroclor-1260-3	6.735	6.634	6.834	546.870	500.000	9.4
Aroclor-1260-4	7.207	7.105	7.305	489.720	500.000	-2.1
Aroclor-1260-5	7.448	7.346	7.546	462.070	500.000	-7.6
Decachlorobiphenyl	8.862	8.759	8.959	43.790	50.000	-12.4
Tetrachloro-m-xylene	3.817	3.716	3.916	47.940	50.000	-4.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 15:29 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.67	5.66	5.56	5.76	-0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.15	6.14	6.04	6.24	-0.01
Aroclor-1260-1	(1)	7.27	7.26	7.16	7.36	0.00
Aroclor-1260-2	(2)	7.52	7.51	7.41	7.61	-0.01
Aroclor-1260-3	(3)	7.88	7.87	7.77	7.97	-0.01
Aroclor-1260-4	(4)	8.10	8.09	7.99	8.19	-0.01
Aroclor-1260-5	(5)	8.42	8.41	8.31	8.51	-0.01
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	-0.01
Decachlorobiphenyl		10.24	10.23	10.13	10.33	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 15:29 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-3	(3)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-4	(4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5	(5)	5.36	5.36	5.26	5.46	0.00
Aroclor-1260-1	(1)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-2	(2)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-3	(3)	6.73	6.73	6.63	6.83	0.00
Aroclor-1260-4	(4)	7.20	7.21	7.11	7.31	0.01
Aroclor-1260-5	(5)	7.45	7.45	7.35	7.55	0.00
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071151.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.671	5.563	5.763	484.620	500.000	-3.1
Aroclor-1016-2	5.692	5.585	5.785	482.920	500.000	-3.4
Aroclor-1016-3	5.755	5.648	5.848	470.770	500.000	-5.8
Aroclor-1016-4	5.852	5.745	5.945	505.800	500.000	1.2
Aroclor-1016-5	6.146	6.038	6.238	510.370	500.000	2.1
Aroclor-1260-1	7.265	7.156	7.356	508.140	500.000	1.6
Aroclor-1260-2	7.519	7.410	7.610	516.880	500.000	3.4
Aroclor-1260-3	7.879	7.768	7.968	504.190	500.000	0.8
Aroclor-1260-4	8.103	7.992	8.192	470.830	500.000	-5.8
Aroclor-1260-5	8.424	8.311	8.511	486.490	500.000	-2.7
Decachlorobiphenyl	10.243	10.125	10.325	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.516	4.412	4.612	50.490	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071151.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.902	4.803	5.003	520.460	500.000	4.1
Aroclor-1016-2	4.921	4.822	5.022	507.900	500.000	1.6
Aroclor-1016-3	5.098	4.999	5.199	543.550	500.000	8.7
Aroclor-1016-4	5.140	5.041	5.241	556.220	500.000	11.2
Aroclor-1016-5	5.355	5.255	5.455	535.660	500.000	7.1
Aroclor-1260-1	6.390	6.292	6.492	521.040	500.000	4.2
Aroclor-1260-2	6.579	6.480	6.680	514.470	500.000	2.9
Aroclor-1260-3	6.732	6.634	6.834	529.510	500.000	5.9
Aroclor-1260-4	7.204	7.105	7.305	494.430	500.000	-1.1
Aroclor-1260-5	7.446	7.346	7.546	448.730	500.000	-10.3
Decachlorobiphenyl	8.859	8.759	8.959	40.800	50.000	-18.4
Tetrachloro-m-xylene	3.816	3.716	3.916	47.690	50.000	-4.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 20:28 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.66	5.56	5.76	-0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.14	6.04	6.24	0.00
Aroclor-1260-1	(1)	7.26	7.26	7.16	7.36	0.00
Aroclor-1260-2	(2)	7.52	7.51	7.41	7.61	-0.01
Aroclor-1260-3	(3)	7.88	7.87	7.77	7.97	-0.01
Aroclor-1260-4	(4)	8.10	8.09	7.99	8.19	-0.01
Aroclor-1260-5	(5)	8.42	8.41	8.31	8.51	-0.01
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	0.00
Decachlorobiphenyl		10.24	10.23	10.13	10.33	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 20:28 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-3	(3)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-4	(4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5	(5)	5.35	5.36	5.26	5.46	0.01
Aroclor-1260-1	(1)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-2	(2)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-3	(3)	6.73	6.73	6.63	6.83	0.00
Aroclor-1260-4	(4)	7.20	7.21	7.11	7.31	0.01
Aroclor-1260-5	(5)	7.44	7.45	7.35	7.55	0.01
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071164.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.669	5.563	5.763	499.770	500.000	0.0
Aroclor-1016-2	5.691	5.585	5.785	486.450	500.000	-2.7
Aroclor-1016-3	5.753	5.648	5.848	487.290	500.000	-2.5
Aroclor-1016-4	5.851	5.745	5.945	522.270	500.000	4.5
Aroclor-1016-5	6.144	6.038	6.238	525.330	500.000	5.1
Aroclor-1260-1	7.263	7.156	7.356	533.550	500.000	6.7
Aroclor-1260-2	7.517	7.410	7.610	536.350	500.000	7.3
Aroclor-1260-3	7.876	7.768	7.968	517.550	500.000	3.5
Aroclor-1260-4	8.100	7.992	8.192	501.150	500.000	0.2
Aroclor-1260-5	8.420	8.311	8.511	506.710	500.000	1.3
Decachlorobiphenyl	10.239	10.125	10.325	50.140	50.000	0.3
Tetrachloro-m-xylene	4.515	4.412	4.612	50.290	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071164.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.901	4.803	5.003	472.560	500.000	-5.5
Aroclor-1016-2	4.919	4.822	5.022	457.460	500.000	-8.5
Aroclor-1016-3	5.096	4.999	5.199	490.060	500.000	-2.0
Aroclor-1016-4	5.139	5.041	5.241	496.520	500.000	-0.7
Aroclor-1016-5	5.353	5.255	5.455	471.190	500.000	-5.8
Aroclor-1260-1	6.389	6.292	6.492	508.230	500.000	1.6
Aroclor-1260-2	6.578	6.480	6.680	502.080	500.000	0.4
Aroclor-1260-3	6.730	6.634	6.834	507.090	500.000	1.4
Aroclor-1260-4	7.202	7.105	7.305	484.720	500.000	-3.1
Aroclor-1260-5	7.444	7.346	7.546	440.870	500.000	-11.8
Decachlorobiphenyl	8.857	8.759	8.959	43.280	50.000	-13.4
Tetrachloro-m-xylene	3.815	3.716	3.916	46.910	50.000	-6.2

Analytical Sequence

Client: Portal Partners Tri-Venture	SDG No.: Q1744
Project: Amtrak Sawtooth Bridges 2025	Instrument ID: ECD_P
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 03/27/2025 03/27/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	03/27/2025	10:14	PP070916.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	03/27/2025	10:30	PP070917.D	10.23	4.51
AR1660ICC750	AR1660ICC750	03/27/2025	10:47	PP070918.D	10.23	4.52
AR1660ICC500	AR1660ICC500	03/27/2025	11:03	PP070919.D	10.23	4.51
AR1660ICC250	AR1660ICC250	03/27/2025	11:19	PP070920.D	10.23	4.52
AR1660ICC050	AR1660ICC050	03/27/2025	11:35	PP070921.D	10.23	4.51
AR1221ICC500	AR1221ICC500	03/27/2025	11:52	PP070922.D	10.23	4.52
AR1232ICC500	AR1232ICC500	03/27/2025	12:11	PP070923.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	03/27/2025	12:28	PP070924.D	10.23	4.51
AR1242ICC750	AR1242ICC750	03/27/2025	12:44	PP070925.D	10.23	4.51
AR1242ICC500	AR1242ICC500	03/27/2025	13:00	PP070926.D	10.23	4.51
AR1242ICC250	AR1242ICC250	03/27/2025	13:16	PP070927.D	10.23	4.51
AR1242ICC050	AR1242ICC050	03/27/2025	13:33	PP070928.D	10.23	4.51
AR1248ICC1000	AR1248ICC1000	03/27/2025	13:49	PP070929.D	10.23	4.51
AR1248ICC750	AR1248ICC750	03/27/2025	14:05	PP070930.D	10.23	4.51
AR1248ICC500	AR1248ICC500	03/27/2025	14:22	PP070931.D	10.23	4.51
AR1248ICC250	AR1248ICC250	03/27/2025	14:38	PP070932.D	10.23	4.52
AR1248ICC050	AR1248ICC050	03/27/2025	14:55	PP070933.D	10.23	4.51
AR1254ICC1000	AR1254ICC1000	03/27/2025	15:11	PP070934.D	10.23	4.51
AR1254ICC750	AR1254ICC750	03/27/2025	15:27	PP070935.D	10.23	4.51
AR1254ICC500	AR1254ICC500	03/27/2025	15:43	PP070936.D	10.23	4.51
AR1254ICC250	AR1254ICC250	03/27/2025	16:00	PP070937.D	10.23	4.51
AR1254ICC050	AR1254ICC050	03/27/2025	16:16	PP070938.D	10.23	4.52
AR1262ICC500	AR1262ICC500	03/27/2025	16:32	PP070939.D	10.23	4.51
AR1268ICC1000	AR1268ICC1000	03/27/2025	16:48	PP070940.D	10.23	4.52
AR1268ICC750	AR1268ICC750	03/27/2025	17:05	PP070941.D	10.23	4.51
AR1268ICC500	AR1268ICC500	03/27/2025	17:21	PP070942.D	10.23	4.51
AR1268ICC250	AR1268ICC250	03/27/2025	17:37	PP070943.D	10.23	4.52
AR1268ICC050	AR1268ICC050	03/27/2025	17:53	PP070944.D	10.23	4.51
AR1660CCC500	AR1660CCC500	04/08/2025	09:37	PP071136.D	10.26	4.52
IBLK	IBLK	04/08/2025	11:05	PP071140.D	10.25	4.52
TR-06-040725MS	Q1742-01MS	04/08/2025	13:41	PP071147.D	10.24	4.51
TR-06-040725MSD	Q1742-01MSD	04/08/2025	13:57	PP071148.D	10.25	4.52
B-158-SB01	Q1744-01	04/08/2025	14:30	PP071150.D	10.24	4.52
AR1660CCC500	AR1660CCC500	04/08/2025	15:29	PP071151.D	10.24	4.52
IBLK	IBLK	04/08/2025	16:51	PP071155.D	10.24	4.52
PB167507BL	PB167507BL	04/08/2025	17:07	PP071156.D	10.24	4.52
PB167507BS	PB167507BS	04/08/2025	17:23	PP071157.D	10.24	4.52
B-158-SB02	Q1744-03	04/08/2025	17:40	PP071158.D	10.24	4.52
AR1660CCC500	AR1660CCC500	04/08/2025	20:28	PP071164.D	10.24	4.52
IBLK	IBLK	04/08/2025	22:05	PP071168.D	10.24	4.52

Analytical Sequence

Analytical Sequence

Client: Portal Partners Tri-Venture	SDG No.: Q1744
Project: Amtrak Sawtooth Bridges 2025	Instrument ID: ECD_P
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 03/27/2025 03/27/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	03/27/2025	10:14	PP070916.D	8.86	3.82
AR1660ICC1000	AR1660ICC1000	03/27/2025	10:30	PP070917.D	8.86	3.82
AR1660ICC750	AR1660ICC750	03/27/2025	10:47	PP070918.D	8.86	3.82
AR1660ICC500	AR1660ICC500	03/27/2025	11:03	PP070919.D	8.86	3.82
AR1660ICC250	AR1660ICC250	03/27/2025	11:19	PP070920.D	8.86	3.82
AR1660ICC050	AR1660ICC050	03/27/2025	11:35	PP070921.D	8.86	3.82
AR1221ICC500	AR1221ICC500	03/27/2025	11:52	PP070922.D	8.86	3.82
AR1232ICC500	AR1232ICC500	03/27/2025	12:11	PP070923.D	8.86	3.82
AR1242ICC1000	AR1242ICC1000	03/27/2025	12:28	PP070924.D	8.86	3.82
AR1242ICC750	AR1242ICC750	03/27/2025	12:44	PP070925.D	8.86	3.82
AR1242ICC500	AR1242ICC500	03/27/2025	13:00	PP070926.D	8.86	3.82
AR1242ICC250	AR1242ICC250	03/27/2025	13:16	PP070927.D	8.86	3.82
AR1242ICC050	AR1242ICC050	03/27/2025	13:33	PP070928.D	8.86	3.82
AR1248ICC1000	AR1248ICC1000	03/27/2025	13:49	PP070929.D	8.86	3.82
AR1248ICC750	AR1248ICC750	03/27/2025	14:05	PP070930.D	8.86	3.82
AR1248ICC500	AR1248ICC500	03/27/2025	14:22	PP070931.D	8.86	3.82
AR1248ICC250	AR1248ICC250	03/27/2025	14:38	PP070932.D	8.86	3.82
AR1248ICC050	AR1248ICC050	03/27/2025	14:55	PP070933.D	8.86	3.82
AR1254ICC1000	AR1254ICC1000	03/27/2025	15:11	PP070934.D	8.86	3.82
AR1254ICC750	AR1254ICC750	03/27/2025	15:27	PP070935.D	8.86	3.82
AR1254ICC500	AR1254ICC500	03/27/2025	15:43	PP070936.D	8.86	3.82
AR1254ICC250	AR1254ICC250	03/27/2025	16:00	PP070937.D	8.86	3.82
AR1254ICC050	AR1254ICC050	03/27/2025	16:16	PP070938.D	8.86	3.82
AR1262ICC500	AR1262ICC500	03/27/2025	16:32	PP070939.D	8.86	3.82
AR1268ICC1000	AR1268ICC1000	03/27/2025	16:48	PP070940.D	8.86	3.82
AR1268ICC750	AR1268ICC750	03/27/2025	17:05	PP070941.D	8.86	3.82
AR1268ICC500	AR1268ICC500	03/27/2025	17:21	PP070942.D	8.86	3.82
AR1268ICC250	AR1268ICC250	03/27/2025	17:37	PP070943.D	8.86	3.82
AR1268ICC050	AR1268ICC050	03/27/2025	17:53	PP070944.D	8.86	3.82
AR1660CCC500	AR1660CCC500	04/08/2025	09:37	PP071136.D	8.86	3.82
IBLK	IBLK	04/08/2025	11:05	PP071140.D	8.86	3.82
TR-06-040725MS	Q1742-01MS	04/08/2025	13:41	PP071147.D	8.86	3.82
TR-06-040725MSD	Q1742-01MSD	04/08/2025	13:57	PP071148.D	8.86	3.82
B-158-SB01	Q1744-01	04/08/2025	14:30	PP071150.D	8.86	3.82
AR1660CCC500	AR1660CCC500	04/08/2025	15:29	PP071151.D	8.86	3.82
IBLK	IBLK	04/08/2025	16:51	PP071155.D	8.86	3.82
PB167507BL	PB167507BL	04/08/2025	17:07	PP071156.D	8.86	3.82
PB167507BS	PB167507BS	04/08/2025	17:23	PP071157.D	8.86	3.82
B-158-SB02	Q1744-03	04/08/2025	17:40	PP071158.D	8.86	3.82
AR1660CCC500	AR1660CCC500	04/08/2025	20:28	PP071164.D	8.86	3.82
IBLK	IBLK	04/08/2025	22:05	PP071168.D	8.86	3.82

Analytical Sequence

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TR-06-040725MS

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Lab Sample ID: Q1742-01MS Date(s) Analyzed: 04/08/2025 04/08/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071147.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.67	5.62	5.72	172	176	4.06
COLUMN 1	2	5.692	5.642	5.742	160		
	3	5.754	5.704	5.804	165		
	4	5.852	5.802	5.902	214		
	5	6.145	6.095	6.195	170		
COLUMN 2	1	4.902	4.852	4.952	167	169	
	2	4.921	4.871	4.971	160		
	3	5.097	5.047	5.147	171		
	4	5.14	5.09	5.19	176		
	5	5.354	5.304	5.404	170		
Aroclor-1260	1	7.265	7.215	7.315	215	188	7.16
COLUMN 1	2	7.519	7.469	7.569	213		
	3	7.878	7.828	7.928	163		
	4	8.103	8.053	8.153	175		
	5	8.423	8.373	8.473	176		
COLUMN 2	1	6.39	6.34	6.44	196	175	
	2	6.579	6.529	6.629	188		
	3	6.732	6.682	6.782	191		
	4	7.204	7.154	7.254	156		
	5	7.445	7.395	7.495	143		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TR-06-040725MSD

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

Lab Sample ID: Q1742-01MSD Date(s) Analyzed: 04/08/2025 04/08/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071148.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.673	5.623	5.723	169	172	0.58
COLUMN 1	2	5.695	5.645	5.745	157		
	3	5.758	5.708	5.808	162		
	4	5.855	5.805	5.905	210		
	5	6.148	6.098	6.198	164		
	1	4.902	4.852	4.952	171		
COLUMN 2	2	4.921	4.871	4.971	164	173	
	3	5.098	5.048	5.148	175		
	4	5.14	5.09	5.19	182		
	5	5.355	5.305	5.405	174		
Aroclor-1260	1	7.269	7.219	7.319	207	177	5.81
COLUMN 1	2	7.522	7.472	7.572	200		
	3	7.881	7.831	7.931	152		
	4	8.106	8.056	8.156	161		
	5	8.426	8.376	8.476	165		
	1	6.391	6.341	6.441	188		
COLUMN 2	2	6.579	6.529	6.629	181	167	
	3	6.733	6.683	6.783	183		
	4	7.205	7.155	7.255	146		
	5	7.445	7.395	7.495	137		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167507BS

Contract: PORT06

Lab Code: CHEM

Case No.: Q1744

SAS No.: Q1744

SDG NO.: Q1744

Lab Sample ID: PB167507BS

Date(s) Analyzed: 04/08/2025

04/08/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071157.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.672	5.622	5.722	178	181	2.8
COLUMN 1	2	5.694	5.644	5.744	179		
	3	5.757	5.707	5.807	178		
	4	5.854	5.804	5.904	189		
	5	6.147	6.097	6.197	184		
	1	4.902	4.852	4.952	170		
COLUMN 2	2	4.92	4.87	4.97	166	176	
	3	5.098	5.048	5.148	176		
	4	5.139	5.089	5.189	189		
	5	5.354	5.304	5.404	179		
Aroclor-1260	1	7.267	7.217	7.317	189	173	8.31
COLUMN 1	2	7.521	7.471	7.571	199		
	3	7.88	7.83	7.93	159		
	4	8.105	8.055	8.155	162		
	5	8.425	8.375	8.475	158		
	1	6.39	6.34	6.44	205	188	
COLUMN 2	2	6.578	6.528	6.628	202		
	3	6.731	6.681	6.781	209		
	4	7.203	7.153	7.253	169		
	5	7.445	7.395	7.495	156		



SAMPLE RAW DATA

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
 Data File : PP071150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2025 14:30
 Operator : YP\AJ
 Sample : Q1744-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 B-158-SB01

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 15:39:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.815	36248933	31402835	24.801	30.293
2) SA Decachlor...	10.244	8.859	24486285	15643655	23.262	19.847

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 14:30
Operator : YP\AJ
Sample : Q1744-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
B-158-SB01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 15:39:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.004 min
Response: 36248933
Conc: 24.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-158-SB01




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 31402835
Conc: 30.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: 0.019 min
Response: 24486285
Conc: 23.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 15643655
Conc: 19.85 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
 Data File : PP071158.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2025 17:40
 Operator : YP\AJ
 Sample : Q1744-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 B-158-SB02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 00:25:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.815	33298055	23846716	22.782	23.004
2) SA Decachlor...	10.241	8.857	17779952	12163887	16.891	15.433

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071158.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 17:40
Operator : YP\AJ
Sample : Q1744-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
B-158-SB02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 00:25:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.003 min
Response: 33298055
Conc: 22.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-158-SB02




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 23846716
Conc: 23.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: 0.016 min
Response: 17779952
Conc: 16.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.857 min
Delta R.T.: -0.002 min
Response: 12163887
Conc: 15.43 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
 Data File : PP071156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2025 17:07
 Operator : YP\AJ
 Sample : PB167507BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167507BL

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 00:24:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.816	38682818	28677374	26.467	27.664
2) SA Decachlor...	10.242	8.858	27306872	18887758	25.942	23.963

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 17:07
Operator : YP\AJ
Sample : PB167507BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167507BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 00:24:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.004 min
Response: 38682818
Conc: 26.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167507BL




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 28677374
Conc: 27.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: 0.016 min
Response: 27306872
Conc: 25.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 18887758
Conc: 23.96 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
 Data File : PP071157.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2025 17:23
 Operator : YP\AJ
 Sample : PB167507BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167507BS

A
B
C
D
E
F
G
H
I
J
K
L

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 00:24:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.518	3.816	38497139	25576966	26.339	24.673
2) SA Decachlor...	10.244	8.857	27062627	17843775	25.710	22.639
Target Compounds						
3) L1 AR-1016-1	5.672	4.902	27097063	20471956	532.869	510.673
4) L1 AR-1016-2	5.694	4.920	41876414	28888897	536.508	499.246
5) L1 AR-1016-3	5.757	5.098	25159092	16299226	533.319	527.086
6) L1 AR-1016-4	5.854	5.139	20816722	13800122	566.353	566.603
7) L1 AR-1016-5	6.147	5.354	18775964	17370284	553.665	537.427
31) L7 AR-1260-1	7.267	6.390	38086561	31565391	567.163	616.435
32) L7 AR-1260-2	7.521	6.578	57469980	38686426	597.796	606.429
33) L7 AR-1260-3	7.880	6.731	38146478	34219624	477.319	628.035 #
34) L7 AR-1260-4	8.105	7.203	38255618	24195511	485.866	505.904
35) L7 AR-1260-5	8.425	7.445	75466317	58288523	474.971	468.142



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071157.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 17:23
Operator : YP\AJ
Sample : PB167507BS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167507BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 00:24:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
 Data File : PP071147.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2025 13:41
 Operator : YP\AJ
 Sample : Q1742-01MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

TR-06-040725MS

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 04/09/2025

Supervised By :mohammad ahmed 04/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 14:15:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.816	27332299	20908590	18.701m	20.170
2) SA Decachlor...	10.243	8.858	24249000	15705126	23.037	19.925
Target Compounds						
3) L1 AR-1016-1	5.670	4.902	22970342	17618377	451.716	439.491
4) L1 AR-1016-2	5.692	4.921	32920730	24304467	421.771	420.020
5) L1 AR-1016-3	5.754	5.097	20545367	13907076	435.518	449.728
6) L1 AR-1016-4	5.852	5.140	20759803	11304852	564.805	464.152
7) L1 AR-1016-5	6.145	5.354	15138267	14442439	446.397	446.841
31) L7 AR-1260-1	7.265	6.390	38056781	26377008	566.719	515.112
32) L7 AR-1260-2	7.519	6.579	53815085	31645634	559.779	496.061
33) L7 AR-1260-3	7.878	6.732	34258366	27370584	428.668	502.334
34) L7 AR-1260-4	8.103	7.204	36215500	19705112	459.956	412.014
35) L7 AR-1260-5	8.423	7.445	73760190	46946118	464.233	377.046

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071147.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 13:41
Operator : YP\AJ
Sample : Q1742-01MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

TR-06-040725MS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 04/09/2025

Supervised By :mohammad ahmed 04/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 14:15:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
 Data File : PP071148.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2025 13:57
 Operator : YP\AJ
 Sample : Q1742-01MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

TR-06-040725MSD

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 04/09/2025

Supervised By :mohammad ahmed 04/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 14:31:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.815	25476701	21963900	17.431m	21.188
2) SA Decachlor...	10.247	8.858	23506426	16173777	22.331	20.520
Target Compounds						
3) L1 AR-1016-1	5.673	4.902	22639729	18047298	445.215	450.190
4) L1 AR-1016-2	5.695	4.921	32357640	24967273	414.556	431.474
5) L1 AR-1016-3	5.758	5.098	20155342	14283830	427.250	461.911
6) L1 AR-1016-4	5.855	5.140	20368243	11652250	554.151	478.416
7) L1 AR-1016-5	6.148	5.355	14646639	14845132	431.900	459.300
31) L7 AR-1260-1	7.269	6.391	36708721	25430138	546.645	496.621
32) L7 AR-1260-2	7.522	6.579	50769132	30372822	528.095	476.109
33) L7 AR-1260-3	7.881	6.733	31997309	26271275	400.376	482.158
34) L7 AR-1260-4	8.106	7.205	33485937	18393580	425.289	384.591
35) L7 AR-1260-5	8.426	7.445	69050583	44825369	434.592	360.013

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040825\
Data File : PP071148.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2025 13:57
Operator : YP\AJ
Sample : Q1742-01MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

TR-06-040725MSD

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 04/09/2025

Supervised By :mohammad ahmed 04/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 14:31:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PP032725	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PP070920.D	AR-1260-3	yogesh	3/28/2025 7:37:36 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC250	PP070920.D	AR-1260-4	yogesh	3/28/2025 7:37:36 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC250	PP070920.D	AR-1260-5	yogesh	3/28/2025 7:37:36 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1016-5	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-1	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-3	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-4	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1660ICC050	PP070921.D	AR-1260-5	yogesh	3/28/2025 7:37:38 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1232ICC500	PP070923.D	Decachlorobiphenyl	yogesh	3/28/2025 7:37:40 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-1	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-2 #2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-3	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software

Manual Integration Report

Sequence:	PP032725	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP070928.D	AR-1242-4	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-4 #2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-5	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1242ICC050	PP070928.D	AR-1242-5 #2	yogesh	3/28/2025 7:37:42 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-2	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-3	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-4	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1248ICC050	PP070933.D	AR-1248-5	yogesh	3/28/2025 7:37:44 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICC1000	PP070934.D	AR-1254-5	yogesh	3/28/2025 7:37:45 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICC750	PP070935.D	AR-1254-5	yogesh	3/28/2025 7:37:47 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICC250	PP070937.D	AR-1254-5	yogesh	3/28/2025 7:37:48 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC1000	PP070940.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:50 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC750	PP070941.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:51 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software

Manual Integration Report

Sequence:	PP032725	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1268ICC500	PP070942.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:53 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC250	PP070943.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:55 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC050	PP070944.D	AR-1268-1 #2	yogesh	3/28/2025 7:37:56 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC050	PP070944.D	AR-1268-4 #2	yogesh	3/28/2025 7:37:56 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1268ICC050	PP070944.D	Tetrachloro-m-xylene	yogesh	3/28/2025 7:37:56 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software
AR1254ICV500	PP070948.D	AR-1254-5	yogesh	3/28/2025 7:37:58 AM	mohammad	3/31/2025 1:17:13	Peak Integrated by Software

Manual Integration Report

Sequence:

PP040825

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP071136.D	AR-1260-4	yogesh	4/9/2025 9:18:11 AM	mohammad	4/10/2025 5:29:57	Peak Integrated by Software
Q1742-01MS	PP071147.D	Tetrachloro-m-xylene	yogesh	4/9/2025 9:18:13 AM	mohammad	4/10/2025 5:29:57	Peak Integrated by Software
Q1742-01MSD	PP071148.D	Tetrachloro-m-xylene	yogesh	4/9/2025 9:18:15 AM	mohammad	4/10/2025 5:29:57	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM
SubDirectory	PP032725	HP Acquire Method	HP Processing Method PP032725
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP070915.D	27 Mar 2025 09:58	YP\AJ	Ok
2	I.BLK	PP070916.D	27 Mar 2025 10:14	YP\AJ	Ok
3	AR1660ICC1000	PP070917.D	27 Mar 2025 10:30	YP\AJ	Ok
4	AR1660ICC750	PP070918.D	27 Mar 2025 10:47	YP\AJ	Ok
5	AR1660ICC500	PP070919.D	27 Mar 2025 11:03	YP\AJ	Ok
6	AR1660ICC250	PP070920.D	27 Mar 2025 11:19	YP\AJ	Ok,M
7	AR1660ICC050	PP070921.D	27 Mar 2025 11:35	YP\AJ	Ok,M
8	AR1221ICC500	PP070922.D	27 Mar 2025 11:52	YP\AJ	Ok
9	AR1232ICC500	PP070923.D	27 Mar 2025 12:11	YP\AJ	Ok,M
10	AR1242ICC1000	PP070924.D	27 Mar 2025 12:28	YP\AJ	Ok
11	AR1242ICC750	PP070925.D	27 Mar 2025 12:44	YP\AJ	Ok
12	AR1242ICC500	PP070926.D	27 Mar 2025 13:00	YP\AJ	Ok
13	AR1242ICC250	PP070927.D	27 Mar 2025 13:16	YP\AJ	Ok
14	AR1242ICC050	PP070928.D	27 Mar 2025 13:33	YP\AJ	Ok,M
15	AR1248ICC1000	PP070929.D	27 Mar 2025 13:49	YP\AJ	Ok
16	AR1248ICC750	PP070930.D	27 Mar 2025 14:05	YP\AJ	Ok
17	AR1248ICC500	PP070931.D	27 Mar 2025 14:22	YP\AJ	Ok
18	AR1248ICC250	PP070932.D	27 Mar 2025 14:38	YP\AJ	Ok
19	AR1248ICC050	PP070933.D	27 Mar 2025 14:55	YP\AJ	Ok,M
20	AR1254ICC1000	PP070934.D	27 Mar 2025 15:11	YP\AJ	Ok,M
21	AR1254ICC750	PP070935.D	27 Mar 2025 15:27	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM
SubDirectory	PP032725	HP Acquire Method	HP Processing Method PP032725
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP070936.D	27 Mar 2025 15:43	YP\AJ	Ok
23	AR1254ICC250	PP070937.D	27 Mar 2025 16:00	YP\AJ	Ok,M
24	AR1254ICC050	PP070938.D	27 Mar 2025 16:16	YP\AJ	Ok
25	AR1262ICC500	PP070939.D	27 Mar 2025 16:32	YP\AJ	Ok
26	AR1268ICC1000	PP070940.D	27 Mar 2025 16:48	YP\AJ	Ok,M
27	AR1268ICC750	PP070941.D	27 Mar 2025 17:05	YP\AJ	Ok,M
28	AR1268ICC500	PP070942.D	27 Mar 2025 17:21	YP\AJ	Ok,M
29	AR1268ICC250	PP070943.D	27 Mar 2025 17:37	YP\AJ	Ok,M
30	AR1268ICC050	PP070944.D	27 Mar 2025 17:53	YP\AJ	Ok,M
31	PP032725ICV500	PP070945.D	27 Mar 2025 18:26	YP\AJ	Ok
32	AR1242ICV500	PP070946.D	27 Mar 2025 18:42	YP\AJ	Ok
33	AR1248ICV500	PP070947.D	27 Mar 2025 19:15	YP\AJ	Ok
34	AR1254ICV500	PP070948.D	27 Mar 2025 19:47	YP\AJ	Ok,M
35	AR1268ICV500	PP070949.D	27 Mar 2025 20:20	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM
SubDirectory	PP040825	HP Acquire Method	HP Processing Method PP032725
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071135.D	08 Apr 2025 08:39	YP\AJ	Ok
2	AR1660CCC500	PP071136.D	08 Apr 2025 09:37	YP\AJ	Ok,M
3	AR1242CCC500	PP071137.D	08 Apr 2025 09:56	YP\AJ	Ok
4	AR1248CCC500	PP071138.D	08 Apr 2025 10:33	YP\AJ	Ok
5	AR1254CCC500	PP071139.D	08 Apr 2025 10:49	YP\AJ	Ok
6	I.BLK	PP071140.D	08 Apr 2025 11:05	YP\AJ	Ok
7	PB167469BL	PP071141.D	08 Apr 2025 11:23	YP\AJ	Not Ok
8	Q1730-15	PP071142.D	08 Apr 2025 11:39	YP\AJ	Ok
9	Q1730-17	PP071143.D	08 Apr 2025 11:56	YP\AJ	Ok
10	Q1730-19	PP071144.D	08 Apr 2025 12:12	YP\AJ	Ok
11	Q1740-01	PP071145.D	08 Apr 2025 13:08	YP\AJ	Ok
12	Q1742-01	PP071146.D	08 Apr 2025 13:25	YP\AJ	Ok,M
13	Q1742-01MS	PP071147.D	08 Apr 2025 13:41	YP\AJ	Ok,M
14	Q1742-01MSD	PP071148.D	08 Apr 2025 13:57	YP\AJ	Ok,M
15	Q1743-01	PP071149.D	08 Apr 2025 14:14	YP\AJ	Ok,M
16	Q1744-01	PP071150.D	08 Apr 2025 14:30	YP\AJ	Ok
17	AR1660CCC500	PP071151.D	08 Apr 2025 15:29	YP\AJ	Ok
18	AR1242CCC500	PP071152.D	08 Apr 2025 16:02	YP\AJ	Ok
19	AR1248CCC500	PP071153.D	08 Apr 2025 16:18	YP\AJ	Ok
20	AR1254CCC500	PP071154.D	08 Apr 2025 16:34	YP\AJ	Ok
21	I.BLK	PP071155.D	08 Apr 2025 16:51	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM
SubDirectory	PP040825	HP Acquire Method	HP Processing Method PP032725
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB167507BL	PP071156.D	08 Apr 2025 17:07	YP\AJ	Ok
23	PB167507BS	PP071157.D	08 Apr 2025 17:23	YP\AJ	Ok
24	Q1744-03	PP071158.D	08 Apr 2025 17:40	YP\AJ	Ok
25	Q1745-01	PP071159.D	08 Apr 2025 17:56	YP\AJ	Ok,M
26	Q1745-09	PP071160.D	08 Apr 2025 18:12	YP\AJ	Ok,M
27	Q1746-01	PP071161.D	08 Apr 2025 18:28	YP\AJ	Ok
28	Q1746-03	PP071162.D	08 Apr 2025 18:45	YP\AJ	Ok
29	Q1747-01	PP071163.D	08 Apr 2025 19:01	YP\AJ	Ok,M
30	AR1660CCC500	PP071164.D	08 Apr 2025 20:28	YP\AJ	Ok
31	AR1242CCC500	PP071165.D	08 Apr 2025 21:17	YP\AJ	Ok
32	AR1248CCC500	PP071166.D	08 Apr 2025 21:33	YP\AJ	Ok
33	AR1254CCC500	PP071167.D	08 Apr 2025 21:49	YP\AJ	Ok
34	I.BLK	PP071168.D	08 Apr 2025 22:05	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM
SubDirectory	PP032725	HP Acquire Method	HP Processing Method PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070915.D	27 Mar 2025 09:58		YP\AJ	Ok
2	I.BLK	I.BLK	PP070916.D	27 Mar 2025 10:14		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP070917.D	27 Mar 2025 10:30		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP070918.D	27 Mar 2025 10:47		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP070919.D	27 Mar 2025 11:03		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP070920.D	27 Mar 2025 11:19		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP070921.D	27 Mar 2025 11:35		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070922.D	27 Mar 2025 11:52		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP070923.D	27 Mar 2025 12:11		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP070924.D	27 Mar 2025 12:28		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP070925.D	27 Mar 2025 12:44		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP070926.D	27 Mar 2025 13:00		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP070927.D	27 Mar 2025 13:16		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP070928.D	27 Mar 2025 13:33		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP070929.D	27 Mar 2025 13:49		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP070930.D	27 Mar 2025 14:05		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP070931.D	27 Mar 2025 14:22		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP070932.D	27 Mar 2025 14:38		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032725

Review By	yogesh	Review On	3/27/2025 12:19:10 PM
Supervise By	mohammad	Supervise On	3/31/2025 1:17:13 AM
SubDirectory	PP032725	HP Acquire Method	HP Processing Method PP032725
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP070933.D	27 Mar 2025 14:55		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP070934.D	27 Mar 2025 15:11		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP070935.D	27 Mar 2025 15:27		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP070936.D	27 Mar 2025 15:43		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP070937.D	27 Mar 2025 16:00		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP070938.D	27 Mar 2025 16:16		YPIAJ	Ok
25	AR1262ICC500	AR1262ICC500	PP070939.D	27 Mar 2025 16:32		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP070940.D	27 Mar 2025 16:48		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP070941.D	27 Mar 2025 17:05		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP070942.D	27 Mar 2025 17:21		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP070943.D	27 Mar 2025 17:37		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP070944.D	27 Mar 2025 17:53		YPIAJ	Ok,M
31	PP032725ICV500	ICVPP032725	PP070945.D	27 Mar 2025 18:26		YPIAJ	Ok
32	AR1242ICV500	ICVPP032725AR1242	PP070946.D	27 Mar 2025 18:42		YPIAJ	Ok
33	AR1248ICV500	ICVPP032725AR1248	PP070947.D	27 Mar 2025 19:15		YPIAJ	Ok
34	AR1254ICV500	ICVPP032725AR1254	PP070948.D	27 Mar 2025 19:47		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP032725AR1268	PP070949.D	27 Mar 2025 20:20		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM
SubDirectory	PP040825	HP Acquire Method	HP Processing Method PP032725

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071135.D	08 Apr 2025 08:39		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071136.D	08 Apr 2025 09:37		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP071137.D	08 Apr 2025 09:56		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071138.D	08 Apr 2025 10:33		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071139.D	08 Apr 2025 10:49		YP\AJ	Ok
6	I.BLK	I.BLK	PP071140.D	08 Apr 2025 11:05		YP\AJ	Ok
7	PB167469BL	PB167469BL	PP071141.D	08 Apr 2025 11:23	not required	YP\AJ	Not Ok
8	Q1730-15	OU4-PCS-TC-25-04032	PP071142.D	08 Apr 2025 11:39		YP\AJ	Ok
9	Q1730-17	OU4-PCS-TC-26-04032	PP071143.D	08 Apr 2025 11:56		YP\AJ	Ok
10	Q1730-19	OU4-CF-15-040325	PP071144.D	08 Apr 2025 12:12	DCB High in 2ND column	YP\AJ	Ok
11	Q1740-01	TP-20	PP071145.D	08 Apr 2025 13:08		YP\AJ	Ok
12	Q1742-01	TR-06-040725	PP071146.D	08 Apr 2025 13:25		YP\AJ	Ok,M
13	Q1742-01MS	TR-06-040725MS	PP071147.D	08 Apr 2025 13:41		YP\AJ	Ok,M
14	Q1742-01MSD	TR-06-040725MSD	PP071148.D	08 Apr 2025 13:57		YP\AJ	Ok,M
15	Q1743-01	TP-16	PP071149.D	08 Apr 2025 14:14	AR1260 Hit	YP\AJ	Ok,M
16	Q1744-01	B-158-SB01	PP071150.D	08 Apr 2025 14:30	TCMX High in 2nd column	YP\AJ	Ok
17	AR1660CCC500	AR1660CCC500	PP071151.D	08 Apr 2025 15:29		YP\AJ	Ok
18	AR1242CCC500	AR1242CCC500	PP071152.D	08 Apr 2025 16:02		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP040825

Review By	yogesh	Review On	4/8/2025 3:06:55 PM
Supervise By	mohammad	Supervise On	4/10/2025 5:29:57 AM
SubDirectory	PP040825	HP Acquire Method	HP Processing Method PP032725
STD. NAME	STD REF.#		
Tune/Reschk	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24332,PP24347,PP24352,PP24357 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP071153.D	08 Apr 2025 16:18		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071154.D	08 Apr 2025 16:34		YPIAJ	Ok
21	I.BLK	I.BLK	PP071155.D	08 Apr 2025 16:51		YPIAJ	Ok
22	PB167507BL	PB167507BL	PP071156.D	08 Apr 2025 17:07		YPIAJ	Ok
23	PB167507BS	PB167507BS	PP071157.D	08 Apr 2025 17:23		YPIAJ	Ok
24	Q1744-03	B-158-SB02	PP071158.D	08 Apr 2025 17:40		YPIAJ	Ok
25	Q1745-01	IB-6A-WC	PP071159.D	08 Apr 2025 17:56		YPIAJ	Ok,M
26	Q1745-09	IB-6.5-WC	PP071160.D	08 Apr 2025 18:12		YPIAJ	Ok,M
27	Q1746-01	B-149-SB01	PP071161.D	08 Apr 2025 18:28		YPIAJ	Ok
28	Q1746-03	B-149-SB02	PP071162.D	08 Apr 2025 18:45		YPIAJ	Ok
29	Q1747-01	ARS20-0008	PP071163.D	08 Apr 2025 19:01		YPIAJ	Ok,M
30	AR1660CCC500	AR1660CCC500	PP071164.D	08 Apr 2025 20:28		YPIAJ	Ok
31	AR1242CCC500	AR1242CCC500	PP071165.D	08 Apr 2025 21:17		YPIAJ	Ok
32	AR1248CCC500	AR1248CCC500	PP071166.D	08 Apr 2025 21:33		YPIAJ	Ok
33	AR1254CCC500	AR1254CCC500	PP071167.D	08 Apr 2025 21:49		YPIAJ	Ok
34	I.BLK	I.BLK	PP071168.D	08 Apr 2025 22:05		YPIAJ	Ok

M : Manual Integration

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Weigh By: EH

Balance check: RJ

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 04/08/2025

Extraction Start Time : 08:11

Extraction End Date : 04/08/2025

Extraction End Time : 11:25

Concentration By: EH

Supervisor By : rajesh

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24328
Surrogate	1.0ML	200 PPB	PP24217
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2601
Baked Na2SO4	N/A	EP2599
Sand	N/A	E2865
Hexane	N/A	E3916
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS721.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: NEVAP-02

Envap Temperature: 40 °C


Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/08/25 11:30	RP (Ext. Lab)	R - Pest/PCBLab
	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 04/08/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167507BL	ABLK507	PCB	30.02	N/A	ritesh	Evelyn	10			U7-1
PB167507BS	ALCS507	PCB	30.01	N/A	ritesh	Evelyn	10			2
Q1740-01	TP-20	PCB	30.05	N/A	ritesh	Evelyn	10	E		3
Q1742-01	TR-06-040725	PCB	30.03	N/A	ritesh	Evelyn	10	E		4
Q1742-01MS	TR-06-040725MS	PCB	30.06	N/A	ritesh	Evelyn	10	E		5
Q1742-01MS D	TR-06-040725MSD	PCB	30.09	N/A	ritesh	Evelyn	10	E		6
Q1743-01	TP-16	PCB	30.04	N/A	ritesh	Evelyn	10	D		U2-1
Q1744-01	B-158-SB01	PCB	30.02	N/A	ritesh	Evelyn	10	E		2
Q1744-03	B-158-SB02	PCB	30.09	N/A	ritesh	Evelyn	10	E		3
Q1745-01	IB-6A-WC	PCB	30.07	N/A	ritesh	Evelyn	10	D		4
Q1745-09	IB-6.5-WC	PCB	30.08	N/A	ritesh	Evelyn	10	D		5
Q1746-01	B-149-SB01	PCB	30.06	N/A	ritesh	Evelyn	10	E		6
Q1746-03	B-149-SB02	PCB	30.02	N/A	ritesh	Evelyn	10	F		U3-1
Q1747-01	ARS20-0008	PCB	30.03	N/A	ritesh	Evelyn	10	E		2

* Extracts relinquished on the same date as received.



161501

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 1740 WorkList ID : 188788 Department : Extraction Date : 04-08-2025 08:08:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1740-01	TP-20	Solid	PCB	Cool 4 deg C	PSEG03	L21	04/04/2025	8082A
Q1742-01	TR-06-040725	Solid	PCB	Cool 4 deg C	PSEG05		04/04/2025	8082A
Q1743-01	TP-16	Solid	PCB	Cool 4 deg C	PSEG03	L33	04/07/2025	8082A
Q1744-01	B-158-SB01	Solid	PCB	Cool 4 deg C	PORT06	L41	04/05/2025	8082A
Q1744-03	B-158-SB02	Solid	PCB	Cool 4 deg C	PORT06	L41	04/05/2025	8082A
Q1745-01	IB-6A-WC	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/07/2025	8082A
Q1745-09	IB-6.5-WC	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/07/2025	8082A
Q1746-01	B-149-SB01	Solid	PCB	Cool 4 deg C	PORT06	L41	04/05/2025	8082A
Q1746-03	B-149-SB02	Solid	PCB	Cool 4 deg C	PORT06	L41	04/06/2025	8082A
Q1747-01	ARS20-0008	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/07/2025	8082A

Date/Time : 04/08/25 8:10
Raw Sample Received by: RJ (Ext 1045)
Raw Sample Relinquished by: CJ Sm

Date/Time : 04/08/25 8:35
Raw Sample Received by: CJ Sm
Raw Sample Relinquished by: RJ (Ext 1045)



LAB CHRONICLE

OrderID:	Q1744	OrderDate:	4/7/2025 11:34:00 AM
Client:	Portal Partners Tri-Venture	Project:	Amtrak Sawtooth Bridges 2025
Contact:	Joseph Krupansky	Location:	L41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1744-01	B-158-SB01	SOIL	PCB	8082A	04/05/25	04/08/25	04/08/25	04/07/25
			TPH GC	8015D		04/08/25	04/08/25	
Q1744-03	B-158-SB02	SOIL	PCB	8082A	04/05/25	04/08/25	04/08/25	04/07/25
			TPH GC	8015D		04/08/25	04/08/25	