

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** Q1744 **SAS No.:** Q1744 **SDG NO.:** Q1744

Continuing Calib Date: 04/10/2025 **Initial Calibration Date(s):** 04/02/2025 04/02/2025

Continuing Calib Time: 04:10 **Initial Calibration Time(s):** 17:32 20:44

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	6.96	6.96	6.86	7.06	0.01
2,4-D	8.04	8.04	7.94	8.14	0.01
2,4,5-TP(Silvex)	8.88	8.89	8.79	8.99	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** Q1744 **SAS No.:** Q1744 **SDG NO.:** Q1744

Continuing Calib Date: 04/10/2025 **Initial Calibration Date(s):** 04/02/2025 04/02/2025

Continuing Calib Time: 04:10 **Initial Calibration Time(s):** 17:32 20:44

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.50	7.48	7.38	7.58	-0.01
2,4-D	8.69	8.67	8.57	8.77	-0.02
2,4,5-TP(Silvex)	9.57	9.55	9.45	9.65	-0.02