

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL02 Date Analyzed: 04/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS029773.D Time Analyzed: 01:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.853	8.788	8.988	800.440	712.500	12.3
2,4-D	8.012	7.942	8.142	758.180	705.000	7.5
2,4-DCAA	6.936	6.863	7.063	810.730	750.000	8.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL02 Date Analyzed: 04/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS029773.D Time Analyzed: 01:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.553	9.451	9.651	813.430	712.500	14.2
2,4-D	8.673	8.571	8.771	749.720	705.000	6.3
2,4-DCAA	7.478	7.375	7.575	812.560	750.000	8.3