

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS029779.D Time Analyzed: 04:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.879	8.788	8.988	715.890	712.500	0.5
2,4-D	8.035	7.942	8.142	682.280	705.000	-3.2
2,4-DCAA	6.955	6.863	7.063	735.610	750.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS029779.D Time Analyzed: 04:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.573	9.451	9.651	730.130	712.500	2.5
2,4-D	8.692	8.571	8.771	669.740	705.000	-5.0
2,4-DCAA	7.495	7.375	7.575	718.420	750.000	-4.2