

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1744

Client: Portal Partners Tri-Venture

Analytical Method: **8081B** **Datafile :** PL095232.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167535BS	gamma-BHC (Lindane)	0.5	0.46	ug/L	92				40 (82)	140 (129)	
	Heptachlor	0.5	0.50	ug/L	100				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.49	ug/L	98				40 (81)	140 (124)	
	Endrin	0.5	0.52	ug/L	105				40 (81)	140 (128)	
	Methoxychlor	0.5	0.52	ug/L	104				40 (78)	140 (108)	