

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG No.: Q1744

Instrument ID: BNA_M Calibration Date/Time: 04/09/2025 09:09

Lab File ID: BM049859.D Init. Calib. Date(s): 04/08/2025 04/08/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:35 20:07

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.178	1.196		1.5	
Benzaldehyde	0.752	0.661		-12.1	
Phenol-d6	1.465	1.508		2.9	
Phenol	1.430	1.460		2.1	20.0
bis(2-Chloroethyl)ether	1.121	1.160		3.5	
2-Chlorophenol	1.207	1.229		1.8	
2-Methylphenol	0.881	0.897		1.8	
2,2-oxybis(1-Chloropropane)	1.250	1.279		2.3	
Acetophenone	0.486	0.499		2.7	
3+4-Methylphenols	1.201	1.214		1.1	
n-Nitroso-di-n-propylamine	0.812	0.846	0.050	4.2	
Nitrobenzene-d5	0.359	0.370		3.1	
Hexachloroethane	0.500	0.504		0.8	
Nitrobenzene	0.346	0.354		2.3	
Isophorone	0.602	0.610		1.3	
2-Nitrophenol	0.183	0.186		1.6	20.0
2,4-Dimethylphenol	0.208	0.214		2.9	
bis(2-Chloroethoxy)methane	0.403	0.417		3.5	
2,4-Dichlorophenol	0.345	0.348		0.9	20.0
Naphthalene	1.028	1.045		1.7	
4-Chloroaniline	0.363	0.379		4.4	
Hexachlorobutadiene	0.247	0.252		2.0	20.0
Caprolactam	0.099	0.097		-2.0	
4-Chloro-3-methylphenol	0.302	0.306		1.3	20.0
2-Methylnaphthalene	0.724	0.748		3.3	
Hexachlorocyclopentadiene	0.245	0.263	0.050	7.3	
2,4,6-Trichlorophenol	0.445	0.457		2.7	20.0
2-Fluorobiphenyl	1.475	1.541		4.5	
2,4,5-Trichlorophenol	0.484	0.490		1.2	
1,1-Biphenyl	1.487	1.543		3.8	
2-Chloronaphthalene	1.169	1.198		2.5	
2-Nitroaniline	0.270	0.281		4.1	
Dimethylphthalate	1.411	1.453		3.0	
Acenaphthylene	1.702	1.748		2.7	
2,6-Dinitrotoluene	0.296	0.310		4.7	
3-Nitroaniline	0.286	0.305		6.6	
Acenaphthene	1.083	1.114		2.9	20.0
2,4-Dinitrophenol	0.175	0.181	0.050	3.4	
4-Nitrophenol	0.255	0.271	0.050	6.3	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG No.: Q1744

Instrument ID: BNA_M Calibration Date/Time: 04/09/2025 09:09

Lab File ID: BM049859.D Init. Calib. Date(s): 04/08/2025 04/08/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:35 20:07

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.810	1.863		2.9	
2,4-Dinitrotoluene	0.394	0.418		6.1	
Diethylphthalate	1.353	1.394		3.0	
4-Chlorophenyl-phenylether	0.771	0.808		4.8	
Fluorene	1.439	1.500		4.2	
4-Nitroaniline	0.295	0.317		7.5	
4,6-Dinitro-2-methylphenol	0.126	0.128		1.6	
n-Nitrosodiphenylamine	0.599	0.620		3.5	20.0
2,4,6-Tribromophenol	0.291	0.294		1.0	
4-Bromophenyl-phenylether	0.232	0.238		2.6	
Hexachlorobenzene	0.266	0.274		3.0	
Atrazine	0.155	0.137		-11.6	
Pentachlorophenol	0.196	0.192		-2.0	20.0
Phenanthrene	1.096	1.136		3.7	
Anthracene	1.081	1.120		3.6	
Carbazole	1.018	1.055		3.6	
Di-n-butylphthalate	1.172	1.208		3.1	
Fluoranthene	1.348	1.391		3.2	20.0
Pyrene	1.378	1.446		4.9	
Terphenyl-d14	1.067	1.223		14.6	
Butylbenzylphthalate	0.518	0.527		1.7	
3,3-Dichlorobenzidine	0.502	0.507		1.0	
Benzo (a) anthracene	1.329	1.372		3.2	
Chrysene	1.264	1.297		2.6	
Bis(2-ethylhexyl)phthalate	0.755	0.791		4.8	
Di-n-octyl phthalate	1.320	1.321		0.1	20.0
Benzo (b) fluoranthene	1.277	1.298		1.6	
Benzo (k) fluoranthene	1.238	1.290		4.2	
Benzo (a) pyrene	1.122	1.138		1.4	20.0
Indeno (1,2,3-cd) pyrene	1.491	1.514		1.5	
Dibenzo (a,h) anthracene	1.228	1.243		1.2	
Benzo (g,h,i) perylene	1.255	1.265		0.8	
1,2,4,5-Tetrachlorobenzene	0.700	0.722		3.1	
1,4-Dioxane	0.485	0.491		1.2	20.0
2,3,4,6-Tetrachlorophenol	0.424	0.427		0.7	

All other compounds must meet a minimum RRF of 0.010.