

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG No.: Q1744

Instrument ID: BNA_M Calibration Date/Time: 04/16/2025 09:28

Lab File ID: BM049944.D Init. Calib. Date(s): 04/08/2025 04/08/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:35 20:07

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.178	1.130		-4.1	
Benzaldehyde	0.752	1.425		89.5	
Phenol-d6	1.465	1.430		-2.4	
Phenol	1.430	1.390		-2.8	20.0
bis(2-Chloroethyl)ether	1.121	1.132		1.0	
2-Chlorophenol	1.207	1.170		-3.1	
2-Methylphenol	0.881	0.869		-1.4	
2,2-oxybis(1-Chloropropane)	1.250	1.250		0.0	
Acetophenone	0.486	0.466		-4.1	
3+4-Methylphenols	1.201	1.188		-1.1	
n-Nitroso-di-n-propylamine	0.812	0.823	0.050	1.4	
Nitrobenzene-d5	0.359	0.349		-2.8	
Hexachloroethane	0.500	0.476		-4.8	
Nitrobenzene	0.346	0.333		-3.8	
Isophorone	0.602	0.598		-0.7	
2-Nitrophenol	0.183	0.184		0.5	20.0
2,4-Dimethylphenol	0.208	0.207		-0.5	
bis(2-Chloroethoxy)methane	0.403	0.396		-1.7	
2,4-Dichlorophenol	0.345	0.338		-2.0	20.0
Naphthalene	1.028	0.990		-3.7	
4-Chloroaniline	0.363	0.343		-5.5	
Hexachlorobutadiene	0.247	0.242		-2.0	20.0
Caprolactam	0.099	0.097		-2.0	
4-Chloro-3-methylphenol	0.302	0.302		0.0	20.0
2-Methylnaphthalene	0.724	0.714		-1.4	
Hexachlorocyclopentadiene	0.245	0.207	0.050	-15.5	
2,4,6-Trichlorophenol	0.445	0.441		-0.9	20.0
2-Fluorobiphenyl	1.475	1.434		-2.8	
2,4,5-Trichlorophenol	0.484	0.481		-0.6	
1,1-Biphenyl	1.487	1.435		-3.5	
2-Chloronaphthalene	1.169	1.113		-4.8	
2-Nitroaniline	0.270	0.275		1.9	
Dimethylphthalate	1.411	1.396		-1.1	
Acenaphthylene	1.702	1.657		-2.6	
2,6-Dinitrotoluene	0.296	0.304		2.7	
3-Nitroaniline	0.286	0.287		0.3	
Acenaphthene	1.083	1.048		-3.2	20.0
2,4-Dinitrophenol	0.175	0.173	0.050	-1.1	
4-Nitrophenol	0.255	0.247	0.050	-3.1	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG No.: Q1744

Instrument ID: BNA_M Calibration Date/Time: 04/16/2025 09:28

Lab File ID: BM049944.D Init. Calib. Date(s): 04/08/2025 04/08/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:35 20:07

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.810	1.768		-2.3	
2,4-Dinitrotoluene	0.394	0.416		5.6	
Diethylphthalate	1.353	1.343		-0.7	
4-Chlorophenyl-phenylether	0.771	0.772		0.1	
Fluorene	1.439	1.410		-2.0	
4-Nitroaniline	0.295	0.296		0.3	
4,6-Dinitro-2-methylphenol	0.126	0.121		-4.0	
n-Nitrosodiphenylamine	0.599	0.573		-4.3	20.0
2,4,6-Tribromophenol	0.291	0.301		3.4	
4-Bromophenyl-phenylether	0.232	0.231		-0.4	
Hexachlorobenzene	0.266	0.264		-0.8	
Atrazine	0.155	0.191		23.2	
Pentachlorophenol	0.196	0.179		-8.7	20.0
Phenanthrene	1.096	1.053		-3.9	
Anthracene	1.081	1.050		-2.9	
Carbazole	1.018	0.972		-4.5	
Di-n-butylphthalate	1.172	1.152		-1.7	
Fluoranthene	1.348	1.334		-1.0	20.0
Pyrene	1.378	1.390		0.9	
Terphenyl-d14	1.067	1.150		7.8	
Butylbenzylphthalate	0.518	0.514		-0.8	
3,3-Dichlorobenzidine	0.502	0.446		-11.2	
Benzo (a) anthracene	1.329	1.294		-2.6	
Chrysene	1.264	1.209		-4.4	
Bis(2-ethylhexyl)phthalate	0.755	0.731		-3.2	
Di-n-octyl phthalate	1.320	1.247		-5.5	20.0
Benzo (b) fluoranthene	1.277	1.237		-3.1	
Benzo (k) fluoranthene	1.238	1.213		-2.0	
Benzo (a) pyrene	1.122	1.079		-3.8	20.0
Indeno (1,2,3-cd) pyrene	1.491	1.332		-10.7	
Dibenzo (a,h) anthracene	1.228	1.092		-11.1	
Benzo (g,h,i) perylene	1.255	1.086		-13.5	
1,2,4,5-Tetrachlorobenzene	0.700	0.678		-3.1	
1,4-Dioxane	0.485	0.447		-7.8	20.0
2,3,4,6-Tetrachlorophenol	0.424	0.416		-1.9	

All other compounds must meet a minimum RRF of 0.010.