

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040925\
 Data File : BM049859.D
 Acq On : 09 Apr 2025 09:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 09 11:48:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM040825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 09 04:00:55 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.485	0.491	-1.2	99	0.00
3	Pyridine	1.274	1.269	0.4	97	0.00
4	n-Nitrosodimethylamine	0.485	0.500	-3.1	100	0.00
5 S	2-Fluorophenol	1.178	1.196	-1.5	99	0.00
6	Aniline	1.256	1.342	-6.8	99	0.00
7 S	Phenol-d6	1.465	1.508	-2.9	99	0.00
8	2-Chlorophenol	1.207	1.229	-1.8	99	0.00
9	Benzaldehyde	0.752	0.661	12.1	86	0.00
10 C	Phenol	1.430	1.460	-2.1	99	0.00
11	bis(2-Chloroethyl)ether	1.121	1.160	-3.5	100	0.00
12	1,3-Dichlorobenzene	1.427	1.453	-1.8	99	0.00
13 C	1,4-Dichlorobenzene	1.436	1.457	-1.5	98	0.00
14	1,2-Dichlorobenzene	1.352	1.379	-2.0	98	0.00
15	Benzyl Alcohol	0.923	0.934	-1.2	96	0.00
16	2,2'-oxybis(1-Chloropropane	1.250	1.279	-2.3	99	0.00
17	2-Methylphenol	0.881	0.897	-1.8	99	0.00
18	Hexachloroethane	0.500	0.504	-0.8	98	0.00
19 P	n-Nitroso-di-n-propylamine	0.812	0.846	-4.2	99	0.00
20	3+4-Methylphenols	1.201	1.214	-1.1	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.486	0.499	-2.7	99	0.00
23 S	Nitrobenzene-d5	0.359	0.370	-3.1	100	0.00
24	Nitrobenzene	0.346	0.354	-2.3	98	0.00
25	Isophorone	0.602	0.610	-1.3	98	0.00
26 C	2-Nitrophenol	0.183	0.186	-1.6	97	0.00
27	2,4-Dimethylphenol	0.208	0.214	-2.9	99	0.00
28	bis(2-Chloroethoxy)methane	0.403	0.417	-3.5	100	0.00
29 C	2,4-Dichlorophenol	0.345	0.348	-0.9	97	0.00
30	1,2,4-Trichlorobenzene	0.399	0.403	-1.0	99	0.00
31	Naphthalene	1.028	1.045	-1.7	99	0.00
32	Benzoic acid	0.220	0.241	-9.5	104	0.00
33	4-Chloroaniline	0.363	0.379	-4.4	99	0.00
34 C	Hexachlorobutadiene	0.247	0.252	-2.0	100	0.00
35	Caprolactam	0.099	0.097	2.0	94	0.00
36 C	4-Chloro-3-methylphenol	0.302	0.306	-1.3	98	0.00
37	2-Methylnaphthalene	0.724	0.748	-3.3	101	0.00
38	1-Methylnaphthalene	0.705	0.724	-2.7	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.700	0.722	-3.1	102	0.00
41 P	Hexachlorocyclopentadiene	0.245	0.263	-7.3	99	0.00
42 S	2,4,6-Tribromophenol	0.291	0.294	-1.0	99	0.00
43 C	2,4,6-Trichlorophenol	0.445	0.457	-2.7	99	0.00
44	2,4,5-Trichlorophenol	0.484	0.490	-1.2	97	0.00
45 S	2-Fluorobiphenyl	1.475	1.541	-4.5	101	0.00
46	1,1'-Biphenyl	1.487	1.543	-3.8	100	0.00
47	2-Chloronaphthalene	1.169	1.198	-2.5	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040925\
 Data File : BM049859.D
 Acq On : 09 Apr 2025 09:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 09 11:48:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM040825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 09 04:00:55 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.270	0.281	-4.1	98	0.00
49	Acenaphthylene	1.702	1.748	-2.7	100	0.00
50	Dimethylphthalate	1.411	1.453	-3.0	100	0.00
51	2,6-Dinitrotoluene	0.296	0.310	-4.7	99	0.00
52 C	Acenaphthene	1.083	1.114	-2.9	99	0.00
53	3-Nitroaniline	0.286	0.305	-6.6	98	0.00
54 P	2,4-Dinitrophenol	0.175	0.181	-3.4	98	0.00
55	Dibenzofuran	1.810	1.863	-2.9	100	0.00
56 P	4-Nitrophenol	0.255	0.271	-6.3	100	0.00
57	2,4-Dinitrotoluene	0.394	0.418	-6.1	99	0.00
58	Fluorene	1.439	1.500	-4.2	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.424	0.427	-0.7	99	0.00
60	Diethylphthalate	1.353	1.394	-3.0	100	0.00
61	4-Chlorophenyl-phenylether	0.771	0.808	-4.8	102	0.00
62	4-Nitroaniline	0.295	0.317	-7.5	98	0.00
63	Azobenzene	1.158	1.210	-4.5	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.126	0.128	-1.6	97	0.00
66 c	n-Nitrosodiphenylamine	0.599	0.620	-3.5	100	0.00
67	4-Bromophenyl-phenylether	0.232	0.238	-2.6	100	0.00
68	Hexachlorobenzene	0.266	0.274	-3.0	100	0.00
69	Atrazine	0.155	0.137	11.6	106	0.00
70 C	Pentachlorophenol	0.196	0.192	2.0	94	0.00
71	Phenanthrene	1.096	1.136	-3.6	100	0.00
72	Anthracene	1.081	1.120	-3.6	100	0.00
73	Carbazole	1.018	1.055	-3.6	100	0.00
74	Di-n-butylphthalate	1.172	1.208	-3.1	99	0.00
75 C	Fluoranthene	1.348	1.391	-3.2	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.265	0.439	-65.7#	107	0.00
78	Pyrene	1.378	1.446	-4.9	101	0.00
79 S	Terphenyl-d14	1.067	1.223	-14.6	101	0.00
80	Butylbenzylphthalate	0.518	0.527	-1.7	97	0.00
81	Benzo(a)anthracene	1.329	1.372	-3.2	99	0.00
82	3,3'-Dichlorobenzidine	0.502	0.507	-1.0	95	0.00
83	Chrysene	1.264	1.297	-2.6	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.755	0.791	-4.8	99	0.00
85 c	Di-n-octyl phthalate	1.320	1.321	-0.1	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	95	-0.01
87	Indeno(1,2,3-cd)pyrene	1.491	1.514	-1.5	96	-0.02
88	Benzo(b)fluoranthene	1.277	1.298	-1.6	94	0.00
89	Benzo(k)fluoranthene	1.238	1.290	-4.2	99	0.00
90 C	Benzo(a)pyrene	1.122	1.138	-1.4	96	-0.01
91	Dibenzo(a,h)anthracene	1.228	1.243	-1.2	95	-0.02
92	Benzo(g,h,i)perylene	1.255	1.265	-0.8	95	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040925\
Data File : BM049859.D
Acq On : 09 Apr 2025 09:09
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Apr 09 11:48:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM040825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 09 04:00:55 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0