

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1744

SDG NO.: Q1744

Lab File ID: BM049858.D

DFTPP Injection Date: 04/09/2025

Instrument ID: BNA_M

DFTPP Injection Time: 08:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.8
68	Less than 2.0% of mass 69	0.4 (1.4) 1
69	Mass 69 relative abundance	28.7
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	35.4
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
275	10.0 - 60.0% of mass 198	25.8
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	11.4
442	Greater than 50% of mass 198	71.5
443	15.0 - 24.0% of mass 442	13.8 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM049859.D	04/09/2025	09:09
B-158-SB01	Q1744-01	BM049864.D	04/09/2025	12:29
B-158-SB01MS	Q1744-01MS	BM049865.D	04/09/2025	13:08
B-158-SB01MSD	Q1744-01MSD	BM049866.D	04/09/2025	13:47
B-158-SB02	Q1744-03	BM049869.D	04/09/2025	15:45