

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1744

SDG NO.: Q1744

Lab File ID: BM049875.D

DFTPP Injection Date: 04/10/2025

Instrument ID: BNA\_M

DFTPP Injection Time: 08:53

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	21.6
68	Less than 2.0% of mass 69	0.4 ( 1.4 ) 1
69	Mass 69 relative abundance	25.7
70	Less than 2.0% of mass 69	0.1 ( 0.5 ) 1
127	10.0 - 80.0% of mass 198	33.2
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.9
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	11.7
442	Greater than 50% of mass 198	76
443	15.0 - 24.0% of mass 442	15 ( 19.7 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM049876.D	04/10/2025	09:32
PB167509BL	PB167509BL	BM049877.D	04/10/2025	10:11