

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG No.: Q1744

Instrument ID: MSVOA_X Calibration Date/Time: 04/08/2025 08:24

Lab File ID: VX045635.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.703	0.685		-2.56	20
1,1-Dichloroethene	0.600	0.591		-1.5	20
2-Butanone	0.550	0.562		2.18	20
Carbon Tetrachloride	0.518	0.541		4.44	20
Chloroform	1.306	1.256		-3.83	20
Benzene	1.463	1.424		-2.67	20
1,2-Dichloroethane	0.604	0.606		0.33	20
Trichloroethene	0.348	0.339		-2.59	20
Tetrachloroethene	0.353	0.350		-0.85	20
Chlorobenzene	1.068	1.057	0.3	-1.03	20
1,2-Dichloroethane-d4	0.914	0.864		-5.47	20
Dibromofluoromethane	0.355	0.347		-2.25	20
Toluene-d8	1.238	1.189		-3.96	20
4-Bromofluorobenzene	0.451	0.456		1.11	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.