

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045524.D	01 Apr 2025 16:15	JC/MD	Ok
2	VSTDICCC001	VX045525.D	01 Apr 2025 17:06	JC/MD	Ok,M
3	VSTDICCC005	VX045526.D	01 Apr 2025 17:29	JC/MD	Ok,M
4	VSTDICCC020	VX045527.D	01 Apr 2025 17:52	JC/MD	Ok,M
5	VSTDICCC050	VX045528.D	01 Apr 2025 18:15	JC/MD	Ok,M
6	VSTDICCC100	VX045529.D	01 Apr 2025 18:38	JC/MD	Ok,M
7	VSTDICCC150	VX045530.D	01 Apr 2025 19:02	JC/MD	Ok,M
8	IBLK	VX045531.D	01 Apr 2025 19:25	JC/MD	Ok
9	VSTDICV050	VX045532.D	01 Apr 2025 19:48	JC/MD	Ok,M
10	BFB	VX045533.D	02 Apr 2025 09:30	JC/MD	Ok
11	VSTDICCC050	VX045534.D	02 Apr 2025 10:02	JC/MD	Ok,M
12	VX0402MBL01	VX045535.D	02 Apr 2025 10:30	JC/MD	Ok
13	VX0402WBL01	VX045536.D	02 Apr 2025 10:53	JC/MD	Ok
14	VX0402WBS01	VX045537.D	02 Apr 2025 11:16	JC/MD	Ok,M
15	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44	JC/MD	Ok,M
16	Q1697-01	VX045539.D	02 Apr 2025 12:07	JC/MD	Dilution
17	Q1697-02	VX045540.D	02 Apr 2025 12:30	JC/MD	Dilution
18	IBLK	VX045541.D	02 Apr 2025 12:54	JC/MD	Ok
19	Q1697-01DL	VX045542.D	02 Apr 2025 13:21	JC/MD	Ok
20	Q1697-02DL	VX045543.D	02 Apr 2025 13:44	JC/MD	Ok
21	Q1697-03	VX045544.D	02 Apr 2025 14:07	JC/MD	Not Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1697-04	VX045545.D	02 Apr 2025 14:31	JC/MD	Not Ok
23	Q1697-05	VX045546.D	02 Apr 2025 14:54	JC/MD	Not Ok
24	Q1697-07	VX045547.D	02 Apr 2025 15:18	JC/MD	Not Ok
25	IBLK	VX045548.D	02 Apr 2025 15:41	JC/MD	Ok
26	Q1697-03	VX045549.D	02 Apr 2025 16:04	JC/MD	Ok
27	Q1697-04	VX045550.D	02 Apr 2025 16:27	JC/MD	Ok
28	Q1697-05	VX045551.D	02 Apr 2025 16:51	JC/MD	Ok
29	Q1697-06	VX045552.D	02 Apr 2025 17:14	JC/MD	ReRun
30	Q1623-01	VX045553.D	02 Apr 2025 17:37	JC/MD	Ok
31	Q1623-02	VX045554.D	02 Apr 2025 18:01	JC/MD	Ok
32	Q1623-03	VX045555.D	02 Apr 2025 18:24	JC/MD	Ok
33	Q1623-04	VX045556.D	02 Apr 2025 18:47	JC/MD	Ok
34	VSTDCCC050	VX045557.D	02 Apr 2025 19:10	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX040825

Review By	John Carlone	Review On	4/9/2025 9:21:19 AM
Supervise By	Maresh Dadoda	Supervise On	4/9/2025 11:41:10 AM
SubDirectory	VX040825	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133619		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133620,VP133621		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045634.D	08 Apr 2025 07:49	JC/MD	Ok
2	VSTDCCC050	VX045635.D	08 Apr 2025 08:24	JC/MD	Ok,M
3	VX0408MBL01	VX045636.D	08 Apr 2025 08:49	JC/MD	Ok
4	VX0408WBL01	VX045637.D	08 Apr 2025 09:12	JC/MD	Ok
5	VX0408WBS01	VX045638.D	08 Apr 2025 09:43	JC/MD	Ok,M
6	VX0408WBSD01	VX045639.D	08 Apr 2025 10:08	JC/MD	Ok,M
7	Q1741-01	VX045640.D	08 Apr 2025 10:32	JC/MD	Ok
8	Q1746-07	VX045641.D	08 Apr 2025 10:55	JC/MD	Ok
9	Q1702-03	VX045642.D	08 Apr 2025 11:18	JC/MD	Ok
10	PB167492ZHE01	VX045643.D	08 Apr 2025 11:41	JC/MD	Ok
11	PB167492ZHE02	VX045644.D	08 Apr 2025 12:05	JC/MD	Ok
12	Q1737-02	VX045645.D	08 Apr 2025 12:28	JC/MD	Ok
13	PB167492ZHE04	VX045646.D	08 Apr 2025 12:51	JC/MD	Ok
14	PB167492ZHE05	VX045647.D	08 Apr 2025 13:14	JC/MD	Ok
15	PB167492ZHE06	VX045648.D	08 Apr 2025 13:38	JC/MD	Ok
16	PB167492ZHE07	VX045649.D	08 Apr 2025 14:01	JC/MD	Ok
17	PB167492ZHE08	VX045650.D	08 Apr 2025 14:24	JC/MD	Ok
18	PB167492ZHE09	VX045651.D	08 Apr 2025 14:47	JC/MD	Ok
19	PB167492ZHE10	VX045652.D	08 Apr 2025 15:11	JC/MD	Ok
20	PB167492ZHE11	VX045653.D	08 Apr 2025 15:34	JC/MD	Ok
21	PB167492ZHE03	VX045654.D	08 Apr 2025 15:58	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX040825

Review By	John Carlone	Review On	4/9/2025 9:21:19 AM
Supervise By	Mahesh Dadoda	Supervise On	4/9/2025 11:41:10 AM
SubDirectory	VX040825	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133619		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133620,VP133621		

22	Q1745-04	VX045655.D	08 Apr 2025 16:21	JC/MD	Ok
23	Q1732-04	VX045656.D	08 Apr 2025 16:44	JC/MD	Ok
24	Q1740-04	VX045657.D	08 Apr 2025 17:07	JC/MD	Ok
25	Q1745-12	VX045658.D	08 Apr 2025 17:31	JC/MD	Ok
26	Q1743-04	VX045659.D	08 Apr 2025 17:54	JC/MD	Ok
27	Q1744-02	VX045660.D	08 Apr 2025 18:17	JC/MD	Ok
28	Q1744-04	VX045661.D	08 Apr 2025 18:40	JC/MD	Ok
29	VSTDCCC050	VX045662.D	08 Apr 2025 19:04	JC/MD	Ok,M

M : Manual Integration