

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG No.: Q1744

Instrument ID: MSVOA_Y Calibration Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) Y Calibration Time(s): 10:08 12:02

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY021656.D			RRF010 = VY021657.D		RRF020 = VY021658.D		RRF050 = VY021659.D		RRF100 = VY021660.D		RRF150 = VY021661.D	
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD					
Dichlorodifluoromethane	0.602	0.577	0.493	0.506	0.485	0.447	0.518	11.4					
Chloromethane	0.876	0.775	0.713	0.692	0.672	0.644	0.729	11.6					
Vinyl Chloride	0.977	0.886	0.810	0.800	0.758	0.714	0.824	11.5					
Bromomethane	0.709	0.613	0.537	0.513	0.494	0.506	0.562	14.9					
Chloroethane	0.644	0.570	0.544	0.512	0.490	0.476	0.539	11.5					
Trichlorofluoromethane	1.212	1.124	1.104	1.013	0.967	0.914	1.056	10.5					
1,1,2-Trichlorotrifluoroethane	0.668	0.601	0.553	0.536	0.510	0.473	0.557	12.4					
1,1-Dichloroethene	0.589	0.564	0.508	0.514	0.496	0.477	0.524	8.2					
Acetone	0.145	0.123	0.102	0.107	0.087	0.102	0.111	18.3					
Carbon Disulfide	1.956	1.828	1.709	1.696	1.623	1.552	1.727	8.4					
Methyl tert-butyl Ether	1.380	1.335	1.247	1.325	1.256	1.292	1.306	3.9					
Methyl Acetate	0.329	0.286	0.269	0.265	0.243	0.257	0.275	11					
Methylene Chloride	0.870	0.694	0.594	0.564	0.518	0.514	0.626	21.8					
trans-1,2-Dichloroethene	0.634	0.608	0.568	0.567	0.551	0.536	0.577	6.3					
1,1-Dichloroethane	1.178	1.091	1.024	1.040	0.992	0.968	1.049	7.3					
Cyclohexane	1.170	1.032	0.912	0.918	0.886	0.818	0.956	13.1					
2-Butanone	0.181	0.166	0.149	0.159	0.140	0.155	0.158	8.9					
Carbon Tetrachloride	0.637	0.594	0.546	0.564	0.547	0.507	0.566	7.9					
cis-1,2-Dichloroethene	0.700	0.658	0.626	0.652	0.631	0.629	0.649	4.3					
Bromochloromethane	0.513	0.470	0.424	0.426	0.420	0.402	0.442	9.3					
Chloroform	1.235	1.141	1.062	1.081	1.026	1.000	1.091	7.9					
1,1,1-Trichloroethane	1.155	1.013	0.959	0.967	0.929	0.893	0.986	9.3					
Methylcyclohexane	0.593	0.599	0.569	0.619	0.612	0.550	0.590	4.5					
Benzene	1.606	1.527	1.429	1.499	1.445	1.371	1.479	5.6					
1,2-Dichloroethane	0.467	0.445	0.399	0.420	0.392	0.379	0.417	8.1					
Trichloroethene	0.412	0.398	0.362	0.374	0.366	0.348	0.377	6.4					
1,2-Dichloropropane	0.383	0.366	0.334	0.353	0.340	0.325	0.350	6.2					
Bromodichloromethane	0.571	0.546	0.504	0.525	0.508	0.484	0.523	6					
4-Methyl-2-Pentanone	0.232	0.240	0.219	0.241	0.222	0.226	0.230	3.9					
Toluene	0.940	0.943	0.893	0.951	0.928	0.885	0.923	3					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1744</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q1744</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q1744</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>03/27/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>10:08</u>
			<u>12:02</u>

LAB FILE ID:		RRF005 = VY021656.D		RRF010 = VY021657.D		RRF020 = VY021658.D		
		RRF050 = VY021659.D		RRF100 = VY021660.D		RRF150 = VY021661.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.456	0.470	0.439	0.486	0.475	0.461	0.465	3.5
cis-1,3-Dichloropropene	0.553	0.558	0.528	0.565	0.549	0.534	0.548	2.6
1,1,2-Trichloroethane	0.280	0.283	0.249	0.261	0.247	0.240	0.260	6.9
2-Hexanone	0.154	0.154	0.144	0.164	0.148	0.154	0.153	4.2
Dibromochloromethane	0.371	0.367	0.340	0.364	0.348	0.337	0.355	4.1
1,2-Dibromoethane	0.267	0.250	0.230	0.254	0.233	0.229	0.244	6.3
Tetrachloroethene	0.534	0.503	0.453	0.449	0.412	0.382	0.455	12.3
Chlorobenzene	1.244	1.178	1.110	1.155	1.116	1.073	1.146	5.3
Ethyl Benzene	1.949	1.916	1.918	2.051	2.023	1.935	1.965	2.9
m/p-Xylenes	0.726	0.753	0.744	0.787	0.774	0.731	0.753	3.2
o-Xylene	0.665	0.670	0.676	0.738	0.725	0.694	0.695	4.4
Styrene	1.064	1.110	1.134	1.232	1.229	1.181	1.158	5.8
Bromoform	0.241	0.239	0.229	0.237	0.225	0.221	0.232	3.4
Isopropylbenzene	3.628	3.701	3.582	3.826	3.843	3.702	3.714	2.8
1,1,2,2-Tetrachloroethane	0.764	0.720	0.616	0.657	0.619	0.631	0.668	9.1
1,3-Dichlorobenzene	1.895	1.774	1.707	1.748	1.728	1.689	1.757	4.2
1,4-Dichlorobenzene	1.876	1.794	1.704	1.741	1.673	1.626	1.736	5.2
1,2-Dichlorobenzene	1.598	1.607	1.496	1.524	1.486	1.465	1.529	3.9
1,2-Dibromo-3-Chloropropane	0.112	0.111	0.097	0.102	0.096	0.102	0.103	6.6
1,2,4-Trichlorobenzene	0.771	0.781	0.787	0.878	0.870	0.889	0.829	6.6
1,2,3-Trichlorobenzene	0.653	0.681	0.673	0.746	0.737	0.753	0.707	6.1
1,2-Dichloroethane-d4	0.612	0.559	0.502	0.548	0.510	0.520	0.542	7.5
Dibromofluoromethane	0.357	0.321	0.290	0.339	0.325	0.313	0.324	7
Toluene-d8	1.277	1.240	1.127	1.306	1.252	1.201	1.234	5.1
4-Bromofluorobenzene	0.447	0.410	0.378	0.439	0.424	0.407	0.417	6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.