

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM				
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP133499					
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508					
CCC		VP133510					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP133509					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBSD01	VY0327SBSD01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133499			
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508			
CCC		VP133510			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133509			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY041025

Review By	Maresh Dadoda	Review On	4/11/2025 1:59:08 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/11/2025 2:00:45 PM
SubDirectory	VY041025	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133628
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133629,VP133630 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021825.D	10 Apr 2025 08:34		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021826.D	10 Apr 2025 09:51		SY/MD	Ok,M
3	VY0410SBL01	VY0410SBL01	VY021827.D	10 Apr 2025 10:28		SY/MD	Ok
4	VY0410SBS01	VY0410SBS01	VY021828.D	10 Apr 2025 10:59		SY/MD	Ok,M
5	VY0410SBSD01	VY0410SBSD01	VY021829.D	10 Apr 2025 11:22		SY/MD	Ok,M
6	Q1747-01RE	ARS20-0008RE	VY021830.D	10 Apr 2025 12:06	vial-B not purged	SY/MD	Not Ok
7	Q1754-01RE	TP-1RE	VY021831.D	10 Apr 2025 12:29	vial-B Internal standard fail;Not match for com.#64	SY/MD	Not Ok
8	Q1732-03	TT-8-VOC	VY021832.D	10 Apr 2025 12:53	vial-A	SY/MD	Ok
9	Q1746-01	B-149-SB01	VY021833.D	10 Apr 2025 13:16	vial-A	SY/MD	Ok
10	Q1746-03	B-149-SB02	VY021834.D	10 Apr 2025 13:40	vial-A	SY/MD	Ok
11	Q1744-01	B-158-SB01	VY021835.D	10 Apr 2025 14:03	vial-A	SY/MD	Ok
12	Q1744-03	B-158-SB02	VY021836.D	10 Apr 2025 14:26	vial-A	SY/MD	Ok
13	Q1738-01RE	OK-02-040425RE	VY021837.D	10 Apr 2025 14:50	vial-B Internal Standard Fail	SY/MD	Confirms
14	Q1749-03	TP-14-VOC	VY021838.D	10 Apr 2025 15:13	vial-A	SY/MD	Ok
15	Q1743-03	TP-16-VOC	VY021839.D	10 Apr 2025 15:37	vial-A	SY/MD	Ok
16	Q1752-01	TP-1	VY021840.D	10 Apr 2025 16:00	vial-A	SY/MD	Ok
17	Q1752-03	TP-2	VY021841.D	10 Apr 2025 16:24	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY041025

Review By	Mahesh Dadoda	Review On	4/11/2025 1:59:08 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	4/11/2025 2:00:45 PM		
SubDirectory	VY041025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133628				
CCC	VP133629,VP133630				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1752-05	TP-4	VY021842.D	10 Apr 2025 16:47	vial-A	SY/MD	Ok
19	Q1752-07	TP-3	VY021843.D	10 Apr 2025 17:10	vial-A	SY/MD	Ok
20	Q1771-03	TP-6C-VOC	VY021844.D	10 Apr 2025 17:34	vial-A	SY/MD	Ok
21	Q1771-07	TP-7C-VOC	VY021845.D	10 Apr 2025 17:57	vial-A	SY/MD	Ok
22	Q1771-11	TP-3A-VOC	VY021846.D	10 Apr 2025 18:21	vial-A	SY/MD	Ok
23	Q1760-03	TP-17-VOC	VY021847.D	10 Apr 2025 18:44	vial-A	SY/MD	Ok
24	Q1760-07	TP-15-VOC	VY021848.D	10 Apr 2025 19:08	vial-A Internal Standard Fail; Surrogate Fail;Need MeOH	SY/MD	Dilution

M : Manual Integration