

Surrogate Summary

SDG No.: Q1744

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1744-01	B-158-SB01	1,2-Dichloroethane-d4	50	57.3	115	70 (63)	130 (155)
		Dibromofluoromethane	50	52.6	105	70 (70)	130 (134)
		Toluene-d8	50	49.1	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	40.5	81	70 (38)	130 (136)
Q1744-03	B-158-SB02	1,2-Dichloroethane-d4	50	55.9	112	70 (63)	130 (155)
		Dibromofluoromethane	50	52.6	105	70 (70)	130 (134)
		Toluene-d8	50	49.1	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	37.6	75	70 (38)	130 (136)
VY0410SBL01	VY0410SBL01	1,2-Dichloroethane-d4	50	46.6	93	70 (63)	130 (155)
		Dibromofluoromethane	50	49.9	100	70 (70)	130 (134)
		Toluene-d8	50	48.5	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	38.3	77	70 (38)	130 (136)
VY0410SBS01	VY0410SBS01	1,2-Dichloroethane-d4	50	49.2	98	70 (63)	130 (155)
		Dibromofluoromethane	50	53.4	107	70 (70)	130 (134)
		Toluene-d8	50	54.3	109	70 (74)	130 (123)
		4-Bromofluorobenzene	50	53.3	107	70 (38)	130 (136)
VY0410SBSD01	VY0410SBSD01	1,2-Dichloroethane-d4	50	48.5	97	70 (63)	130 (155)
		Dibromofluoromethane	50	50.9	102	70 (70)	130 (134)
		Toluene-d8	50	51.4	103	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.7	101	70 (38)	130 (136)

