

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744
 Lab File ID: VY021826.D Date Analyzed: 04/10/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	208417	7.71	315600	8.62	284701	11.42
UPPER LIMIT	416834	8.213	631200	9.122	569402	11.92
LOWER LIMIT	104209	7.213	157800	8.122	142351	10.92
EPA SAMPLE NO.						
B-158-SB01	219582	7.71	409739	8.62	351379	11.42
B-158-SB02	221830	7.71	407665	8.62	345604	11.42
VY0410SBL01	241941	7.71	436661	8.62	362482	11.42
VY0410SBS01	202932	7.71	310477	8.62	276303	11.42
VY0410SBSD01	207811	7.71	318832	8.62	284393	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.