

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1744 SAS No.: Q1744 SDG NO.: Q1744
Lab File ID: VY021825.D BFB Injection Date: 04/10/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:34
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.3
75	30.0 - 60.0% of mass 95	54.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.2
173	Less than 2.0% of mass 174	1.4 (1.7) 1
174	50.0 - 100.0% of mass 95	81.7
175	5.0 - 9.0% of mass 174	6.7 (8.2) 1
176	95.0 - 101.0% of mass 174	78.2 (95.7) 1
177	5.0 - 9.0% of mass 176	5.1 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021826.D	04/10/2025	09:51
VY0410SBL01	VY0410SBL01	VY021827.D	04/10/2025	10:28
VY0410SBS01	VY0410SBS01	VY021828.D	04/10/2025	10:59
VY0410SBSD01	VY0410SBSD01	VY021829.D	04/10/2025	11:22
B-158-SB01	Q1744-01	VY021835.D	04/10/2025	14:03
B-158-SB02	Q1744-03	VY021836.D	04/10/2025	14:26