

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2025

Project # N/A

Chemtech Project # Q1746

Test Name: EPH

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 04/07/2025.

4 Water samples were received on 04/07/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TPH GC, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for EPH.

C. Analytical Techniques:

The analysis were performed on instrument FID_C. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analyses were performed on instrument FID_D. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis of EPHs was based on method NJEPH and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS {Q1746-01MS} with File ID: FD049276.D recoveries met the requirements for all compounds except for Aromatic C10-C12[34.9%] Due to matrix interference.

The MSD {Q1746-01MSD} with File ID: FD049277.D recoveries met the acceptable requirements except for Aromatic C10-C12[35%] Due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration met the requirements .

Sample B-158-GW01 was diluted due to high concentration for aliphatic compound .

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____