



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q1746
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1746 **SAS No.:** Q1746
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	86	40 - 160	CV	04/09/2025	11:47	LB135354
CRI01	Aluminum	98.0	100	98	40 - 160	P	04/10/2025	10:33	LB135373
	Antimony	51.0	50.0	102	40 - 160	P	04/10/2025	10:33	LB135373
	Arsenic	18.9	20.0	95	40 - 160	P	04/10/2025	10:33	LB135373
	Barium	85.4	100	85	40 - 160	P	04/10/2025	10:33	LB135373
	Beryllium	5.48	6.0	91	40 - 160	P	04/10/2025	10:33	LB135373
	Cadmium	5.51	6.0	92	40 - 160	P	04/10/2025	10:33	LB135373
	Calcium	1880	2000	94	40 - 160	P	04/10/2025	10:33	LB135373
	Chromium	9.42	10.0	94	40 - 160	P	04/10/2025	10:33	LB135373
	Cobalt	28.2	30.0	94	40 - 160	P	04/10/2025	10:33	LB135373
	Copper	21.9	20.0	109	40 - 160	P	04/10/2025	10:33	LB135373
	Iron	102	100	102	40 - 160	P	04/10/2025	10:33	LB135373
	Lead	11.6	12.0	97	40 - 160	P	04/10/2025	10:33	LB135373
	Magnesium	1980	2000	99	40 - 160	P	04/10/2025	10:33	LB135373
	Manganese	15.0	20.0	75	40 - 160	P	04/10/2025	10:33	LB135373
	Nickel	38.0	40.0	95	40 - 160	P	04/10/2025	10:33	LB135373
	Potassium	1960	2000	98	40 - 160	P	04/10/2025	10:33	LB135373
	Selenium	18.1	20.0	91	40 - 160	P	04/10/2025	10:33	LB135373
	Silver	10.4	10.0	104	40 - 160	P	04/10/2025	10:33	LB135373
	Sodium	1480	2000	74	40 - 160	P	04/10/2025	10:33	LB135373
	Thallium	38.3	40.0	96	40 - 160	P	04/10/2025	10:33	LB135373
	Vanadium	36.3	40.0	91	40 - 160	P	04/10/2025	10:33	LB135373
	Zinc	39.6	40.0	99	40 - 160	P	04/10/2025	10:33	LB135373
CRA	Mercury	0.21	0.2	104	40 - 160	CV	04/11/2025	10:12	LB135388
CRI01	Aluminum	91.7	100	92	40 - 160	P	04/14/2025	14:09	LB135424
	Antimony	51.7	50.0	103	40 - 160	P	04/14/2025	14:09	LB135424
	Arsenic	22.3	20.0	112	40 - 160	P	04/14/2025	14:09	LB135424
	Barium	95.6	100	96	40 - 160	P	04/14/2025	14:09	LB135424
	Beryllium	5.76	6.0	96	40 - 160	P	04/14/2025	14:09	LB135424
	Cadmium	6.08	6.0	101	40 - 160	P	04/14/2025	14:09	LB135424
	Calcium	1970	2000	98	40 - 160	P	04/14/2025	14:09	LB135424
	Chromium	9.67	10.0	97	40 - 160	P	04/14/2025	14:09	LB135424
	Cobalt	29.3	30.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Copper	22.7	20.0	113	40 - 160	P	04/14/2025	14:09	LB135424
	Iron	102	100	102	40 - 160	P	04/14/2025	14:09	LB135424

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q1746

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1746 **SAS No.:** Q1746

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	12.2	12.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Magnesium	2060	2000	103	40 - 160	P	04/14/2025	14:09	LB135424
	Manganese	20.4	20.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Nickel	39.4	40.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Potassium	1930	2000	97	40 - 160	P	04/14/2025	14:09	LB135424
	Selenium	19.6	20.0	98	40 - 160	P	04/14/2025	14:09	LB135424
	Silver	9.88	10.0	99	40 - 160	P	04/14/2025	14:09	LB135424
	Sodium	1840	2000	92	40 - 160	P	04/14/2025	14:09	LB135424
	Thallium	41.8	40.0	104	40 - 160	P	04/14/2025	14:09	LB135424
	Vanadium	40.7	40.0	102	40 - 160	P	04/14/2025	14:09	LB135424
	Zinc	41.0	40.0	102	40 - 160	P	04/14/2025	14:09	LB135424