

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Portal Partners Tri-Venture</u>	SDG No.: <u>Q1746</u>
Contract: <u>PORT06</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q1746</u>
	SAS No.: <u>Q1746</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.16			-20	20	04/10/2025	10:37	LB135373
	Barium	-3.92	6.0	65	-94	106	04/10/2025	10:37	LB135373
	Cadmium	-3.53	1.0	353	-5	7	04/10/2025	10:37	LB135373
	Chromium	55.8	52.0	107	42	62	04/10/2025	10:37	LB135373
	Lead	7.32			-12	12	04/10/2025	10:37	LB135373
	Selenium	-8.19			-20	20	04/10/2025	10:37	LB135373
	Silver	3.75			-10	10	04/10/2025	10:37	LB135373
ICSAB01	Arsenic	111	104	107	88.4	120	04/10/2025	10:42	LB135373
	Barium	459	537	86	437	637	04/10/2025	10:42	LB135373
	Cadmium	981	972	101	826	1120	04/10/2025	10:42	LB135373
	Chromium	553	542	102	460	624	04/10/2025	10:42	LB135373
	Lead	57.3	49.0	117	37	61	04/10/2025	10:42	LB135373
	Selenium	36.7	46.0	80	26	66	04/10/2025	10:42	LB135373
	Silver	186	201	92	170	232	04/10/2025	10:42	LB135373
ICSA01	Arsenic	10.8			-20	20	04/14/2025	14:14	LB135424
	Barium	4.69	6.0	78	-94	106	04/14/2025	14:14	LB135424
	Cadmium	-1.94	1.0	194	-5	7	04/14/2025	14:14	LB135424
	Chromium	55.2	52.0	106	42	62	04/14/2025	14:14	LB135424
	Lead	5.03			-12	12	04/14/2025	14:14	LB135424
	Selenium	-0.72			-20	20	04/14/2025	14:14	LB135424
	Silver	-3.74			-10	10	04/14/2025	14:14	LB135424
ICSAB01	Arsenic	114	104	110	88.4	120	04/14/2025	14:18	LB135424
	Barium	479	537	89	437	637	04/14/2025	14:18	LB135424
	Cadmium	1040	972	107	826	1120	04/14/2025	14:18	LB135424
	Chromium	553	542	102	460	624	04/14/2025	14:18	LB135424
	Lead	53.8	49.0	110	37	61	04/14/2025	14:18	LB135424
	Selenium	47.1	46.0	102	26	66	04/14/2025	14:18	LB135424
	Silver	179	201	89	170	232	04/14/2025	14:18	LB135424