

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** Q1746
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1746 **SAS No.:** Q1746
Matrix: Water **Sample ID:** Q1743-04 **Client ID:** TP-16DUP
Percent Solids for Sample: NA **Duplicate ID** Q1743-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	2470		2440		1		P
Cadmium	ug/L	20	4.92	J	5.01	J	2		P
Chromium	ug/L	20	50.0	U	14.2	J	200.0		P
Lead	ug/L	20	2220		2220		0		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** Q1746
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1746 **SAS No.:** Q1746
Matrix: Water **Sample ID:** Q1743-04MS **Client ID:** TP-16MSD
Percent Solids for Sample: NA **Duplicate ID** Q1743-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3390		3500		3		P
Barium	ug/L	20	3120		3020		3		P
Cadmium	ug/L	20	865		872		1		P
Chromium	ug/L	20	1680		1700		1		P
Lead	ug/L	20	6040		6110		1		P
Mercury	ug/L	20	36.6		35.5		3		CV
Selenium	ug/L	20	8320		8540		3		P
Silver	ug/L	20	310		314		1		P