

Continuing Calibration Report for SequenceID : FC041125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068546.D

Aliphatic C9-C12	6138470.000	60.000	3.345	6.644	102307.833	105282.003	2.825
Aliphatic C12-C16	3913351.000	40.000	6.645	10.043	97833.775	102322.575	4.387
Aliphatic C16-C21	5581584.000	60.000	10.044	13.409	93026.400	98816.539	5.859
Aliphatic C21-C28	7034588.000	80.000	13.410	17.073	87932.350	94185.337	6.639
Aliphatic C28-C40	10265171.000	120.000	17.074	22.075	85543.092	89882.835	4.828
Aliphatic EPH	32933164.000	360.000	3.345	22.075	91481.011	96276.618	4.981