

Continuing Calibration Report for SequenceID : FD040925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049265.D

Aromatic C10-C12	5823092.000	40.000	4.440	6.185	145577.300	156792.089	7.153
Aromatic C12-C16	9287317.000	60.000	6.186	8.813	154788.617	170225.083	9.068
Aromatic C16-C21	12670535.000	80.000	8.814	13.103	158381.688	167701.469	5.557
Aromatic C21-C36	25214080.000	180.000	13.104	18.518	140078.222	138140.429	-1.403
Aromatic EPH	52995024.000	360.000	4.440	18.518	147208.400	152129.398	3.235