

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1746 **SAS No :** Q1746 **SDG No:** Q1746
Sample No : PB167524BSD **Datafile:** FC068538.D
Client ID : PB167524BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	300	166	55		0.6 (40-140)	50
Aliphatic C12-C16	200	144	72		2.1 (40-140)	50
Aliphatic C16-C21	300	248	83		3.1 (40-140)	50
Aliphatic C21-C28	400	335	84		2.7 (40-140)	50
Aliphatic C28-C40	600	514	86		0.737 (40-140)	50