

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1746 **SAS No :** Q1746 **SDG No:** Q1746
Sample No : PB167513BSD **Datafile:** FC068562.D
Client ID : PB167513BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10.0	5.89	59		2.8 (40-140)	50
Aliphatic C12-C16	6.7	5.49	82		1.1 (40-140)	50
Aliphatic C16-C21	10.0	9.61	96		1 (40-140)	50
Aliphatic C21-C28	13.3	12.9	97		1.6 (40-140)	50
Aliphatic C28-C40	20.0	19.4	97		1.5 (40-140)	50