

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1746 **SAS No :** Q1746 **SDG No:** Q1746
Sample No : PB167524BSD **Datafile:** FD049260.D
Client ID : PB167524BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aromatic C10-C12	200	105	52		2.8 (40-140)	50
Aromatic C12-C16	300	184	61		0.54 (40-140)	50
Aromatic C16-C21	500	401	80		3.3 (40-140)	50
Aromatic C21-C36	800	717	90		1.2 (40-140)	50