

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 10:33 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.68	5.67	5.57	5.77	-0.01
Aroclor-1248-2 (2)	5.95	5.94	5.84	6.04	-0.01
Aroclor-1248-3 (3)	6.15	6.14	6.04	6.24	-0.01
Aroclor-1248-4 (4)	6.55	6.54	6.44	6.64	-0.01
Aroclor-1248-5 (5)	6.59	6.58	6.48	6.68	-0.01
Tetrachloro-m-xylene	4.52	4.51	4.41	4.61	-0.01
Decachlorobiphenyl	10.25	10.23	10.13	10.33	-0.02

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 10:33 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1248-2	(2)	5.14	5.14	5.04	5.24	0.00
Aroclor-1248-3	(3)	5.18	5.18	5.08	5.28	0.00
Aroclor-1248-4	(4)	5.36	5.36	5.26	5.46	0.00
Aroclor-1248-5	(5)	5.75	5.75	5.65	5.85	0.00
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00