

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 10:49 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.52	6.52	6.42	6.62	0.00
Aroclor-1254-2	(2)	6.74	6.73	6.63	6.83	-0.01
Aroclor-1254-3	(3)	7.10	7.10	7.00	7.20	0.00
Aroclor-1254-4	(4)	7.38	7.38	7.28	7.48	0.00
Aroclor-1254-5	(5)	7.80	7.79	7.69	7.89	-0.01
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	0.00
Decachlorobiphenyl		10.24	10.23	10.13	10.33	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 10:49 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.71	5.71	5.61	5.81	0.00
Aroclor-1254-2 (2)	5.86	5.86	5.76	5.96	0.00
Aroclor-1254-3 (3)	6.26	6.26	6.16	6.36	0.00
Aroclor-1254-4 (4)	6.49	6.49	6.39	6.59	0.00
Aroclor-1254-5 (5)	6.91	6.91	6.81	7.01	0.00
Tetrachloro-m-xylene	3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl	8.86	8.86	8.76	8.96	0.00