

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071151.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.671	5.563	5.763	484.620	500.000	-3.1
Aroclor-1016-2	5.692	5.585	5.785	482.920	500.000	-3.4
Aroclor-1016-3	5.755	5.648	5.848	470.770	500.000	-5.8
Aroclor-1016-4	5.852	5.745	5.945	505.800	500.000	1.2
Aroclor-1016-5	6.146	6.038	6.238	510.370	500.000	2.1
Aroclor-1260-1	7.265	7.156	7.356	508.140	500.000	1.6
Aroclor-1260-2	7.519	7.410	7.610	516.880	500.000	3.4
Aroclor-1260-3	7.879	7.768	7.968	504.190	500.000	0.8
Aroclor-1260-4	8.103	7.992	8.192	470.830	500.000	-5.8
Aroclor-1260-5	8.424	8.311	8.511	486.490	500.000	-2.7
Decachlorobiphenyl	10.243	10.125	10.325	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.516	4.412	4.612	50.490	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071151.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.803	5.003	520.460	500.000	4.1
Aroclor-1016-2	4.921	4.822	5.022	507.900	500.000	1.6
Aroclor-1016-3	5.098	4.999	5.199	543.550	500.000	8.7
Aroclor-1016-4	5.140	5.041	5.241	556.220	500.000	11.2
Aroclor-1016-5	5.355	5.255	5.455	535.660	500.000	7.1
Aroclor-1260-1	6.390	6.292	6.492	521.040	500.000	4.2
Aroclor-1260-2	6.579	6.480	6.680	514.470	500.000	2.9
Aroclor-1260-3	6.732	6.634	6.834	529.510	500.000	5.9
Aroclor-1260-4	7.204	7.105	7.305	494.430	500.000	-1.1
Aroclor-1260-5	7.446	7.346	7.546	448.730	500.000	-10.3
Decachlorobiphenyl	8.859	8.759	8.959	40.800	50.000	-18.4
Tetrachloro-m-xylene	3.816	3.716	3.916	47.690	50.000	-4.6