

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL05 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071164.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.669	5.563	5.763	499.770	500.000	0.0
Aroclor-1016-2	5.691	5.585	5.785	486.450	500.000	-2.7
Aroclor-1016-3	5.753	5.648	5.848	487.290	500.000	-2.5
Aroclor-1016-4	5.851	5.745	5.945	522.270	500.000	4.5
Aroclor-1016-5	6.144	6.038	6.238	525.330	500.000	5.1
Aroclor-1260-1	7.263	7.156	7.356	533.550	500.000	6.7
Aroclor-1260-2	7.517	7.410	7.610	536.350	500.000	7.3
Aroclor-1260-3	7.876	7.768	7.968	517.550	500.000	3.5
Aroclor-1260-4	8.100	7.992	8.192	501.150	500.000	0.2
Aroclor-1260-5	8.420	8.311	8.511	506.710	500.000	1.3
Decachlorobiphenyl	10.239	10.125	10.325	50.140	50.000	0.3
Tetrachloro-m-xylene	4.515	4.412	4.612	50.290	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL05 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PP071164.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.803	5.003	472.560	500.000	-5.5
Aroclor-1016-2	4.919	4.822	5.022	457.460	500.000	-8.5
Aroclor-1016-3	5.096	4.999	5.199	490.060	500.000	-2.0
Aroclor-1016-4	5.139	5.041	5.241	496.520	500.000	-0.7
Aroclor-1016-5	5.353	5.255	5.455	471.190	500.000	-5.8
Aroclor-1260-1	6.389	6.292	6.492	508.230	500.000	1.6
Aroclor-1260-2	6.578	6.480	6.680	502.080	500.000	0.4
Aroclor-1260-3	6.730	6.634	6.834	507.090	500.000	1.4
Aroclor-1260-4	7.202	7.105	7.305	484.720	500.000	-3.1
Aroclor-1260-5	7.444	7.346	7.546	440.870	500.000	-11.8
Decachlorobiphenyl	8.857	8.759	8.959	43.280	50.000	-13.4
Tetrachloro-m-xylene	3.815	3.716	3.916	46.910	50.000	-6.2