

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL01 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029764.D Time Analyzed: 20:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	8.863	8.788	8.988	735.080	712.500	3.2
2,4-D	8.020	7.942	8.142	723.340	705.000	2.6
2,4-DCAA	6.943	6.863	7.063	740.470	750.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL01 Date Analyzed: 04/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS029764.D Time Analyzed: 20:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.538	9.451	9.651	765.850	712.500	7.5
2,4-D	8.659	8.571	8.771	705.850	705.000	0.1
2,4-DCAA	7.467	7.375	7.575	764.180	750.000	1.9