

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

Continuing Calib Date: 04/15/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 11:21 Initial Calibration Time(s): 15:07 16:15

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.06       | 9.05      | 8.95                   | 9.15 | -0.01      |
| Tetrachloro-m-xylene | 3.54       | 3.54      | 3.44                   | 3.64 | 0.00       |
| gamma-BHC (Lindane)  | 4.33       | 4.32      | 4.22                   | 4.42 | -0.01      |
| Heptachlor           | 4.92       | 4.91      | 4.81                   | 5.01 | 0.00       |
| Heptachlor epoxide   | 5.68       | 5.68      | 5.58                   | 5.78 | 0.00       |
| Endrin               | 6.57       | 6.57      | 6.47                   | 6.67 | 0.00       |
| Methoxychlor         | 7.50       | 7.50      | 7.40                   | 7.60 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

Continuing Calib Date: 04/15/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 11:21 Initial Calibration Time(s): 15:07 16:15

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Decachlorobiphenyl   | 7.90       | 7.90      | 7.80      | 8.00 | 0.00       |
| Tetrachloro-m-xylene | 2.77       | 2.77      | 2.67      | 2.87 | 0.00       |
| gamma-BHC (Lindane)  | 3.60       | 3.60      | 3.50      | 3.70 | 0.00       |
| Heptachlor           | 3.94       | 3.94      | 3.84      | 4.04 | 0.00       |
| Heptachlor epoxide   | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Endrin               | 5.63       | 5.63      | 5.53      | 5.73 | 0.00       |
| Methoxychlor         | 6.60       | 6.60      | 6.50      | 6.70 | 0.00       |