

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 04/09/2025

Lab Sample No.: PSTDCCC050 Data File : PL095138.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.048	8.956	9.156	48.170	50.000	-3.7
Endrin	6.567	6.474	6.674	42.670	50.000	-14.7
gamma-BHC (Lindane)	4.323	4.227	4.427	45.190	50.000	-9.6
Heptachlor	4.910	4.815	5.015	42.960	50.000	-14.1
Heptachlor epoxide	5.678	5.583	5.783	44.200	50.000	-11.6
Methoxychlor	7.495	7.400	7.600	48.130	50.000	-3.7
Tetrachloro-m-xylene	3.534	3.438	3.638	46.390	50.000	-7.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 04/09/2025

Lab Sample No.: PSTDCCC050 Data File : PL095138.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.899	7.807	8.007	40.930	50.000	-18.1
Endrin	5.626	5.534	5.734	46.910	50.000	-6.2
gamma-BHC (Lindane)	3.599	3.504	3.704	50.840	50.000	1.7
Heptachlor	3.936	3.842	4.042	47.670	50.000	-4.7
Heptachlor epoxide	4.717	4.625	4.825	47.170	50.000	-5.7
Methoxychlor	6.599	6.507	6.707	44.680	50.000	-10.6
Tetrachloro-m-xylene	2.768	2.672	2.872	51.110	50.000	2.2