

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 04/09/2025

Lab Sample No.: PSTDCCC050 Data File : PL095149.D Time Analyzed: 20:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.047	8.956	9.156	50.060	50.000	0.1
Endrin	6.567	6.474	6.674	40.440	50.000	-19.1
gamma-BHC (Lindane)	4.322	4.227	4.427	47.980	50.000	-4.0
Heptachlor	4.909	4.815	5.015	44.270	50.000	-11.5
Heptachlor epoxide	5.677	5.583	5.783	46.130	50.000	-7.7
Methoxychlor	7.492	7.400	7.600	46.330	50.000	-7.3
Tetrachloro-m-xylene	3.534	3.438	3.638	49.030	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 04/09/2025

Lab Sample No.: PSTDCCC050 Data File : PL095149.D Time Analyzed: 20:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.898	7.807	8.007	48.140	50.000	-3.7
Endrin	5.626	5.534	5.734	45.520	50.000	-9.0
gamma-BHC (Lindane)	3.598	3.504	3.704	53.340	50.000	6.7
Heptachlor	3.935	3.842	4.042	49.330	50.000	-1.3
Heptachlor epoxide	4.717	4.625	4.825	50.590	50.000	1.2
Methoxychlor	6.599	6.507	6.707	45.780	50.000	-8.4
Tetrachloro-m-xylene	2.767	2.672	2.872	53.560	50.000	7.1