

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **PORT06**

Lab Code: **CHEM** Case No.: **Q1746** SAS No.: **Q1746** SDG NO.: **Q1746**

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **03/11/2025** **03/11/2025**

Client Sample No. (PEM): **PEM - PL095128.D** Date Analyzed: **04/09/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **08:43**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.049	8.950	9.150	26.150	20.000	30.8
Tetrachloro-m-xylene	3.536	3.490	3.590	23.490	20.000	17.5
alpha-BHC	3.992	3.940	4.040	12.010	10.000	20.1
beta-BHC	4.523	4.470	4.570	12.040	10.000	20.4
gamma-BHC (Lindane)	4.324	4.270	4.370	12.060	10.000	20.6
Endrin	6.569	6.500	6.640	50.710	50.000	1.4
4,4'-DDT	7.020	6.950	7.090	112.470	100.000	12.5
Methoxychlor	7.497	7.430	7.570	271.750	250.000	8.7

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **03/11/2025** **03/11/2025**

Client Sample No. (PEM): **PEM - PL095128.D** Date Analyzed: **04/09/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **08:43**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.900	7.800	8.000	24.640	20.000	23.2
Tetrachloro-m-xylene	2.770	2.720	2.820	24.610	20.000	23.1
alpha-BHC	3.271	3.220	3.320	11.740	10.000	17.4
beta-BHC	3.901	3.850	3.950	13.350	10.000	33.5
gamma-BHC (Lindane)	3.600	3.550	3.650	11.860	10.000	18.6
Endrin	5.627	5.560	5.700	56.090	50.000	12.2
4,4'-DDT	6.026	5.960	6.100	122.140	100.000	22.1
Methoxychlor	6.599	6.530	6.670	276.490	250.000	10.6