

Surrogate Summary

SW-846

SDG No.: Q1746

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167509BL	PB167509BL	2-Fluorophenol	150	138	92		30 (18)	130 (112)
		Phenol-d6	150	134	89		30 (15)	130 (107)
		Nitrobenzene-d5	100	95.5	95		30 (18)	130 (107)
		2-Fluorobiphenyl	100	96.3	96		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	146	97		30 (10)	130 (116)
		Terphenyl-d14	100	112	112		30 (10)	130 (105)
PB167509BS	PB167509BS	2-Fluorophenol	150	127	85		30 (18)	130 (112)
		Phenol-d6	150	126	84		30 (15)	130 (107)
		Nitrobenzene-d5	100	87.2	87		30 (18)	130 (107)
		2-Fluorobiphenyl	100	89.2	89		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	145	97		30 (10)	130 (116)
		Terphenyl-d14	100	101	101		30 (10)	130 (105)
Q1744-01MS	B-158-SB01MS	2-Fluorophenol	150	123	82		30 (18)	130 (112)
		Phenol-d6	150	125	83		30 (15)	130 (107)
		Nitrobenzene-d5	100	82.9	83		30 (18)	130 (107)
		2-Fluorobiphenyl	100	81.7	82		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	135	90		30 (10)	130 (116)
		Terphenyl-d14	100	89.4	89		30 (10)	130 (105)
Q1744-01MSD	B-158-SB01MSD	2-Fluorophenol	150	127	85		30 (18)	130 (112)
		Phenol-d6	150	130	87		30 (15)	130 (107)
		Nitrobenzene-d5	100	86.2	86		30 (18)	130 (107)
		2-Fluorobiphenyl	100	85.2	85		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	141	94		30 (10)	130 (116)
		Terphenyl-d14	100	93.3	93		30 (10)	130 (105)
Q1746-01	B-149-SB01	2-Fluorophenol	150	107	71		30 (18)	130 (112)
		Phenol-d6	150	105	70		30 (15)	130 (107)
		Nitrobenzene-d5	100	73.2	73		30 (18)	130 (107)
		2-Fluorobiphenyl	100	65.6	66		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	103	69		30 (10)	130 (116)
		Terphenyl-d14	100	62.9	63		30 (10)	130 (105)
Q1746-03	B-149-SB02	2-Fluorophenol	150	152	102		30 (18)	130 (112)
		Phenol-d6	150	155	104		30 (15)	130 (107)
		Nitrobenzene-d5	100	105	105		30 (18)	130 (107)
		2-Fluorobiphenyl	100	101	101		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	177	118		30 (10)	130 (116)
		Terphenyl-d14	100	109	109		30 (10)	130 (105)

Surrogate Summary

SW-846

SDG No.: Q1746

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167521BL	PB167521BL	2-Fluorophenol	150	128	86		15 (10)	110 (139)
		Phenol-d6	150	125	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	88.3	88		30 (49)	130 (133)
		2-Fluorobiphenyl	100	89.2	89		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	136	91		15 (44)	110 (137)
		Terphenyl-d14	100	108	108		30 (48)	130 (125)
PB167521BS	PB167521BS	2-Fluorophenol	150	119	79		15 (10)	110 (139)
		Phenol-d6	150	118	78		15 (10)	110 (134)
		Nitrobenzene-d5	100	79.5	79		30 (49)	130 (133)
		2-Fluorobiphenyl	100	82.6	83		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	130	87		15 (44)	110 (137)
		Terphenyl-d14	100	94.4	94		30 (48)	130 (125)
PB167521BSD	PB167521BSD	2-Fluorophenol	150	125	83		15 (10)	110 (139)
		Phenol-d6	150	122	82		15 (10)	110 (134)
		Nitrobenzene-d5	100	84.5	84		30 (49)	130 (133)
		2-Fluorobiphenyl	100	84.4	84		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	136	91		15 (44)	110 (137)
		Terphenyl-d14	100	98.6	99		30 (48)	130 (125)
Q1746-05	B-158-GW01	2-Fluorophenol	150	37.1	25		15 (10)	110 (139)
		Phenol-d6	150	23.0	15		15 (10)	110 (134)
		Nitrobenzene-d5	100	65.7	66		30 (49)	130 (133)
		2-Fluorobiphenyl	100	61.4	61		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	83.8	56		15 (44)	110 (137)
		Terphenyl-d14	100	58.4	58		30 (48)	130 (125)
Q1746-06	B-149-GW01	2-Fluorophenol	150	68.5	46		15 (10)	110 (139)
		Phenol-d6	150	40.8	27		15 (10)	110 (134)
		Nitrobenzene-d5	100	99.7	100		30 (49)	130 (133)
		2-Fluorobiphenyl	100	94.1	94		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	172	114	*	15 (44)	110 (137)
		Terphenyl-d14	100	96.3	96		30 (48)	130 (125)
Q1746-07	EB-2025-4-7	2-Fluorophenol	150	68.3	46		15 (10)	110 (139)
		Phenol-d6	150	38.6	26		15 (10)	110 (134)
		Nitrobenzene-d5	100	105	105		30 (49)	130 (133)
		2-Fluorobiphenyl	100	92.4	92		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	172	115	*	15 (44)	110 (137)
		Terphenyl-d14	100	104	104		30 (48)	130 (125)